
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003149
14:56 14/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:35
STERILIZING 3:31
136.4 C 223kPa 3:31
136.2 C 222kPa 4:01
136.1 C 217kPa 4:31
136.3 C 223kPa 5:01
136.4 C 225kPa 5:31
136.4 C 223kPa 6:01
136.4 C 223kPa 6:31
136.2 C 222kPa 7:01
136.2 C 222kPa 7:01
VENTING 7:01
AIR DRYING 7:51
CYCLE COMPLETE 7:59

Digital Signature #
4CD1BC3EFC001AFB
