

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2641

Name: RITTER 1  
Sterilizer ID: V2178198  
12/05/2022 05:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.8 F 0.0  
2:00 161.2 F 2.4  
4:00 173.9 F 5.7  
6:00 197.1 F 10.3  
8:00 217.4 F 5.3  
10:00 227.0 F 9.3  
12:00 234.6 F 9.6  
14:00 240.8 F 11.9  
16:00 239.3 F 10.3  
18:00 253.2 F 17.1  
20:00 265.0 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 272.6 F 29.3  
2:00 271.5 F 28.7  
3:00 272.1 F 28.9  
4:00 271.7 F 29.0  
  
Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 228.4 F 0.0  
10:00 237.8 F 0.0  
15:00 229.7 F 0.0  
20:00 224.5 F 0.0  
25:00 223.4 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 21:39  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

12/05/2022 06:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_