
Midmark M11 - v1.0.0
Total Cycles: 3377

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2023 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 80.3 F 0.0
4:00 91.3 F 0.0
6:00 115.5 F 0.0
8:00 143.4 F 1.0
10:00 171.2 F 3.9
12:00 212.7 F 6.9
14:00 220.8 F 10.4
16:00 233.4 F 8.4
18:00 229.2 F 7.6
20:00 238.4 F 11.7
22:00 233.8 F 8.1
24:00 245.3 F 12.8
26:00 243.5 F 12.0
28:00 255.4 F 18.1
30:00 266.0 F 24.5
32:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.7 F 29.1
2:00 271.9 F 28.9
3:00 272.6 F 29.0
4:00 271.9 F 29.0

Min 271.7 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 200.4 F 0.0
10:00 229.2 F 0.0
15:00 227.6 F 0.0
20:00 229.8 F 0.0
25:00 229.1 F 0.0
30:00 230.7 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:30
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

06/13/2023 08:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____