
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000084
10:58 21/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:15
PRESSURIZING 1:23
STERILIZING 2:49
136.4 C 226kPa 2:49
136.3 C 223kPa 3:19
136.4 C 225kPa 3:49
136.2 C 223kPa 4:19
136.2 C 224kPa 4:49
136.4 C 225kPa 5:19
136.0 C 224kPa 5:49
136.5 C 226kPa 6:19
136.5 C 226kPa 6:19
VENTING 6:19
AIR DRYING 7:14
CYCLE COMPLETE 22:14

Digital Signature #
3E646F38225023A4
