
Midmark M11 - v1.0.5
Total Cycles: 2882

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.1 F 0.0
2:00 173.9 F 3.0
4:00 188.5 F 7.5
6:00 209.5 F 13.0
8:00 225.2 F 8.3
10:00 237.7 F 11.5
12:00 240.8 F 11.9
14:00 240.1 F 10.7
16:00 253.4 F 17.4
18:00 264.9 F 24.3
20:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 273.2 F 29.7
2:00 271.6 F 29.5
3:00 272.6 F 29.6
4:00 271.5 F 29.4

Min 271.3 F 29.1
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.8 F 0.0
5:00 209.1 F 0.0
10:00 202.3 F 0.0
15:00 199.3 F 0.0
20:00 197.8 F 0.0
25:00 196.7 F 0.0
30:00 194.6 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

09/14/2023 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____