
Midmark M11 - v1.0.0
Total Cycles: 3342

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 188.8 F 2.2
4:00 205.7 F 7.5
6:00 225.3 F 13.0
8:00 231.1 F 8.8
10:00 240.0 F 10.2
12:00 243.1 F 12.4
14:00 241.6 F 11.1
16:00 254.9 F 17.9
18:00 266.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 28.9
2:00 271.6 F 28.3
3:00 272.8 F 29.1
4:00 272.1 F 28.4

Min 271.6 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 220.3 F 0.0
10:00 225.8 F 0.0
15:00 220.8 F 0.0
20:00 217.6 F 0.0
25:00 217.3 F 0.0
30:00 213.5 F 0.0

Drying Complete

FILLING: 1:25
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

06/06/2023 12:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____