

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000329  
12:55 11/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:56  
STERILIZING 4:13  
136.4 C 233kPa 4:13  
136.4 C 233kPa 4:43  
136.4 C 233kPa 5:13  
136.4 C 233kPa 5:43  
136.2 C 230kPa 6:13  
136.3 C 232kPa 6:43  
136.4 C 233kPa 7:13  
136.4 C 233kPa 7:43  
136.4 C 233kPa 7:43  
VENTING 7:43  
AIR DRYING 8:38  
CYCLE COMPLETE 8:47

-----  
Digital Signature #  
5872DA40CF9D18A0  
-----