
Midmark M11 - v1.0.0
Total Cycles: 2912

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.1 F 0.0
2:00 177.7 F 3.2
4:00 180.7 F 7.0
6:00 207.8 F 11.6
8:00 214.7 F 5.5
10:00 229.5 F 10.6
12:00 230.3 F 6.7
14:00 243.3 F 13.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 240F
CHAMBER PRESS 10.6PSI
** ITEMS NOT STERILE **

02/16/2023 10:38 AM