

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3691

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2024 02:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 151.9 F 2.6  
4:00 169.1 F 6.3  
6:00 191.4 F 10.6  
8:00 212.7 F 5.7  
10:00 222.7 F 9.2  
12:00 232.2 F 10.7  
14:00 238.1 F 11.4  
16:00 236.4 F 9.7  
18:00 249.9 F 15.9  
20:00 260.8 F 22.2  
22:00 270.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.6 F 30.0  
2:00 271.4 F 29.2  
3:00 272.4 F 30.1  
4:00 271.8 F 29.5

Min 271.2 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 228.9 F 0.0  
10:00 211.5 F 0.0  
15:00 213.8 F 0.0  
20:00 221.2 F 0.0  
25:00 226.2 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 22:49  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:01:47  
CYCLE COMPLETE

06/12/2024 03:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_