
Midmark M11 - v1.0.5
Total Cycles: 4237

Name: RITTER2
Sterilizer ID: V2313228
01/02/2025 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 69.3 F 0.0
4:00 74.2 F 0.0
6:00 98.4 F 0.1
8:00 131.7 F 2.0
10:00 151.4 F 4.3
12:00 173.9 F 6.9
14:00 187.7 F 9.8
16:00 205.6 F 13.3
18:00 213.8 F 6.7
20:00 223.8 F 10.6
22:00 230.7 F 8.4
24:00 237.9 F 11.7
26:00 236.7 F 9.7
28:00 248.8 F 15.5
30:00 259.2 F 21.3
32:00 268.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.5 F 30.2
2:00 271.5 F 29.3
3:00 272.5 F 30.1
4:00 271.4 F 29.3

Min 271.1 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 202.9 F 0.0
10:00 202.4 F 0.0
15:00 202.0 F 0.0
20:00 217.5 F 0.0
25:00 223.8 F 0.0
30:00 226.3 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:06
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:11:38
CYCLE COMPLETE

01/02/2025 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____