

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000315  
11:57 27/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:08  
STERILIZING 4:12  
136.0 C 229kPa 4:12  
136.3 C 233kPa 4:42  
136.4 C 233kPa 5:12  
136.2 C 231kPa 5:42  
136.1 C 232kPa 6:12  
136.3 C 232kPa 6:42  
136.4 C 233kPa 7:12  
136.3 C 233kPa 7:42  
136.3 C 233kPa 7:42  
VENTING 7:43  
AIR DRYING 8:37  
CYCLE COMPLETE 23:37  
-----

Digital Signature #  
FECF8FF3DC7E16B0  
-----