

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
3.3uS / 2.1ppm  
CYCLE NUMBER 000007  
8:35 13/01/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:54  
PRESSURIZING 8:04  
STERILIZING 9:55  
136.0 C 222kPa 9:55  
136.4 C 227kPa 10:25  
136.4 C 227kPa 10:55  
136.4 C 226kPa 11:25  
136.3 C 226kPa 11:55  
136.4 C 227kPa 12:25  
136.3 C 226kPa 12:55  
136.4 C 226kPa 13:25  
VENTING 13:25  
AIR DRYING 14:41  
CYCLE COMPLETE 44:41  
-----

Digital Signature #  
73D0C9FB20895105  
-----