
Midmark M11 - v1.0.5
Total Cycles: 3554

Name: RITTER2
Sterilizer ID: V2313228
04/24/2024 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.1 F 0.0
2:00 156.6 F 2.9
4:00 178.0 F 7.0
6:00 202.2 F 12.4
8:00 221.1 F 7.5
10:00 240.3 F 13.4
12:00 237.4 F 10.9
14:00 236.4 F 9.4
16:00 250.8 F 16.4
18:00 262.8 F 23.4
20:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 273.1 F 30.1
2:00 271.6 F 29.7
3:00 272.7 F 30.0
4:00 271.5 F 29.6

Min 271.4 F 29.4
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 251.1 F 0.0
10:00 188.8 F 0.0
15:00 186.8 F 0.0
20:00 184.8 F 0.0
25:00 183.1 F 0.0
30:00 180.9 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

04/24/2024 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____