
Midmark M11 - v1.0.0
Total Cycles: 4950

Name: RITTER 1
Sterilizer ID: V2178198
10/01/2024 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 180.6 F 2.8
4:00 195.8 F 7.3
6:00 215.1 F 12.3
8:00 224.4 F 6.6
10:00 237.9 F 11.6
12:00 235.1 F 8.4
14:00 241.4 F 10.8
16:00 246.3 F 13.1
18:00 258.3 F 19.7
20:00 269.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.5 F 28.0
3:00 272.4 F 29.0
4:00 271.9 F 28.4

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 202.6 F 0.0
10:00 198.2 F 0.0
15:00 201.2 F 0.0
20:00 203.3 F 0.0
25:00 211.1 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

10/01/2024 05:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____