
Midmark M11 - v1.0.0
Total Cycles: 5452

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 153.4 F 3.4
4:00 176.5 F 7.4
6:00 199.4 F 11.8
8:00 218.4 F 5.1
10:00 231.3 F 10.0
12:00 231.1 F 7.8
14:00 241.2 F 11.7
16:00 238.6 F 9.7
18:00 252.0 F 16.2
20:00 263.5 F 23.0
22:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.6
2:00 272.4 F 29.0
3:00 271.6 F 28.2
4:00 272.3 F 28.9

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 247.1 F 0.0
10:00 222.0 F 0.0
15:00 224.3 F 0.0
20:00 224.6 F 0.0
25:00 231.5 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:05
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

03/26/2025 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____