
Midmark M11 - v1.0.0
Total Cycles: 1510

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2022 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.5 F 0.0
2:00 164.0 F 3.1
4:00 177.9 F 7.1
6:00 196.6 F 11.9
8:00 216.2 F 5.3
10:00 225.4 F 10.6
12:00 233.0 F 7.0
14:00 242.1 F 13.1
16:00 242.2 F 11.5
18:00 254.9 F 18.2
20:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 29.1
2:00 272.4 F 29.5
3:00 272.0 F 28.7
4:00 272.7 F 29.5

Min 271.6 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 210.6 F 0.0
10:00 227.3 F 0.0
15:00 232.0 F 0.0
20:00 229.7 F 0.0
25:00 227.3 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

01/11/2022 01:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____