

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1368

Name: RITTER2  
Sterilizer ID: V2313228  
08/04/2022 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
2:00 184.4 F 0.2  
4:00 186.3 F 3.5  
6:00 208.1 F 8.7  
8:00 222.8 F 10.1  
10:00 232.9 F 10.5  
12:00 233.2 F 8.4  
14:00 239.7 F 10.3  
16:00 248.2 F 14.7  
18:00 260.6 F 21.7  
20:00 271.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.4 F 29.6  
2:00 271.5 F 28.9  
3:00 272.0 F 29.7  
4:00 272.3 F 29.4

Min 271.3 F 28.8  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 233.4 F 0.0  
10:00 228.6 F 0.0  
15:00 230.2 F 0.0  
20:00 226.6 F 0.0  
25:00 227.5 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 20:37  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

08/04/2022 11:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_