
Midmark M11 - v1.0.0
Total Cycles: 2247

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 162.1 F 0.3
4:00 186.0 F 0.9
6:00 208.6 F 3.5
8:00 222.9 F 8.2
10:00 236.0 F 13.1
12:00 235.5 F 9.7
14:00 237.0 F 9.1
16:00 245.7 F 12.7
18:00 245.6 F 12.8
20:00 258.3 F 19.6
22:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.5
2:00 271.8 F 28.3
3:00 272.5 F 28.5
4:00 271.8 F 28.3

Min 271.5 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 223.5 F 0.0
10:00 229.6 F 0.0
15:00 232.4 F 0.0
20:00 226.7 F 0.0
25:00 229.0 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:60
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

08/25/2022 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____