
Midmark M11 - v1.0.5
Total Cycles: 709

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 04:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.2 F 0.0
2:00 194.4 F 4.9
4:00 218.6 F 11.8
6:00 231.0 F 8.9
8:00 237.3 F 9.6
10:00 241.3 F 11.1
12:00 249.0 F 15.0
14:00 262.6 F 22.8

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.7
1:00 272.3 F 29.1
2:00 272.7 F 29.5
3:00 271.9 F 29.1
4:00 272.0 F 28.9

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 233.1 F 0.0
10:00 210.9 F 0.0
15:00 213.7 F 0.0
20:00 211.9 F 0.0
25:00 205.6 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 2:54
HEATING: 15:54
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 00:55:48
CYCLE COMPLETE

01/19/2022 05:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____