
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001762
15:50 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:30
STERILIZING 3:39
136.4 C 224kPa 3:39
136.3 C 225kPa 4:09
136.3 C 225kPa 4:39
136.3 C 225kPa 5:09
136.4 C 226kPa 5:39
136.3 C 223kPa 6:09
136.4 C 224kPa 6:39
136.3 C 224kPa 7:09
136.3 C 224kPa 7:09
VENTING 7:09
AIR DRYING 7:56
CYCLE COMPLETE 14:35

Digital Signature #
CDD6399A31F03024
