

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1750

Name: RITTER2  
Sterilizer ID: V2313228  
11/21/2022 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
2:00 195.7 F 2.2  
4:00 206.4 F 8.1  
6:00 224.8 F 10.2  
8:00 234.9 F 11.0  
10:00 236.7 F 9.7  
12:00 237.1 F 9.0  
14:00 252.4 F 16.9  
16:00 265.2 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.3 F 29.2  
2:00 272.1 F 29.5  
3:00 272.2 F 29.1  
4:00 271.9 F 29.4  
  
Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 226.4 F 0.0  
10:00 232.5 F 0.0  
15:00 229.9 F 0.0  
20:00 224.3 F 0.0  
25:00 223.7 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 17:40  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:56:54  
CYCLE COMPLETE

11/21/2022 12:30 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_