
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000220
13:58 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:34
STERILIZING 5:10
136.0 C 229kPa 5:10
136.4 C 233kPa 5:40
136.2 C 232kPa 6:10
136.3 C 233kPa 6:40
136.4 C 233kPa 7:10
136.4 C 233kPa 7:40
136.3 C 232kPa 8:10
136.4 C 233kPa 8:40
136.4 C 233kPa 8:40
VENTING 8:40
AIR DRYING 9:43
CYCLE COMPLETE 23:16

Digital Signature #
74D6089EF1606C11
