

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004723  
14:29 15/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:33  
STERILIZING 3:23  
136.0 C 220kPa 3:23  
136.2 C 222kPa 3:53  
136.3 C 224kPa 4:23  
136.3 C 224kPa 4:53  
136.4 C 223kPa 5:23  
136.4 C 223kPa 5:53  
136.4 C 223kPa 6:23  
136.4 C 224kPa 6:53  
136.4 C 224kPa 6:53  
VENTING 6:53  
AIR DRYING 7:42  
CYCLE COMPLETE 7:49

-----  
Digital Signature #  
BE6918696A1881FB  
-----