
Midmark M11 - v1.0.0
Total Cycles: 1544

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 175.5 F 2.8
4:00 186.6 F 6.8
6:00 203.4 F 11.8
8:00 220.2 F 5.7
10:00 229.5 F 10.6
12:00 231.4 F 7.1
14:00 242.9 F 12.8
16:00 241.7 F 11.1
18:00 254.6 F 17.7
20:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.3 F 29.0
2:00 271.8 F 28.3
3:00 272.3 F 29.0
4:00 272.1 F 28.5

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 234.8 F 0.0
10:00 226.0 F 0.0
15:00 226.9 F 0.0
20:00 229.0 F 0.0
25:00 227.3 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

01/20/2022 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____