
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001008
11:09 17/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:46
136.5 C 224kPa 3:46
136.2 C 224kPa 4:16
136.1 C 217kPa 4:46
136.1 C 222kPa 5:16
136.3 C 225kPa 5:46
136.4 C 225kPa 6:16
136.4 C 223kPa 6:46
136.4 C 220kPa 7:16
136.4 C 220kPa 7:16
VENTING 7:16
AIR DRYING 8:08
CYCLE COMPLETE 8:12

Digital Signature #
E256B8CB081BAC17
