
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001135
7:10 08/02/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:23
STERILIZING 4:31
136.0 C 220kPa 4:31
136.4 C 224kPa 5:01
136.3 C 220kPa 5:31
136.4 C 221kPa 6:01
136.4 C 222kPa 6:31
135.9 C 220kPa 7:01
135.8 C 219kPa 7:31
136.1 C 219kPa 8:01
136.1 C 219kPa 8:01
VENTING 8:01
AIR DRYING 8:50
CYCLE COMPLETE 8:56

Digital Signature #
68540C59F6A4B79B
