

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000907  
15:23 04/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:34  
STERILIZING 4:46  
136.0 C 228kPa 4:46  
136.4 C 233kPa 5:16  
136.3 C 232kPa 5:46  
136.4 C 233kPa 6:16  
136.4 C 233kPa 6:46  
136.3 C 232kPa 7:16  
136.2 C 231kPa 7:46  
136.5 C 233kPa 8:16  
136.5 C 233kPa 8:16  
VENTING 8:16  
AIR DRYING 9:11  
CYCLE COMPLETE 9:27  
-----

Digital Signature #  
FD1CE5B4365E9FF5  
-----