

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1622

Name: RITTER 1  
Sterilizer ID: V2178198  
02/10/2022 07:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.3 F 0.0  
2:00 115.5 F 0.0  
4:00 139.0 F 2.3  
6:00 164.1 F 5.5  
8:00 208.5 F 9.6  
10:00 218.9 F 6.3  
12:00 222.3 F 7.3  
14:00 235.0 F 11.9  
16:00 234.5 F 8.4  
18:00 242.6 F 11.6  
20:00 244.4 F 12.4  
22:00 256.5 F 18.8  
24:00 267.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.5 F 29.2  
2:00 272.0 F 28.6  
3:00 271.3 F 28.5  
4:00 272.8 F 29.3

Min 271.1 F 28.2  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 227.8 F 0.0  
10:00 224.9 F 0.0  
15:00 231.9 F 0.0  
20:00 223.5 F 0.0  
25:00 230.1 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 25:29  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:03:53  
CYCLE COMPLETE

02/10/2022 08:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_