

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007524  
10:50 10/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:29  
STERILIZING 3:19  
136.0 C 220kPa 3:19  
136.4 C 224kPa 3:49  
136.4 C 224kPa 4:19  
136.2 C 222kPa 4:49  
136.2 C 218kPa 5:19  
136.2 C 224kPa 5:49  
136.4 C 225kPa 6:19  
136.0 C 220kPa 6:49  
136.0 C 220kPa 6:49  
VENTING 6:49  
AIR DRYING 7:44  
CYCLE COMPLETE 7:52

-----  
Digital Signature #  
C102CA8441454E3A  
-----