

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2713

Name: RITTER2  
Sterilizer ID: V2313228  
07/31/2023 02:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.7 F 0.0  
2:00 133.5 F 0.0  
4:00 143.7 F 0.0  
6:00 183.4 F 0.0  
8:00 203.8 F 1.3  
10:00 217.1 F 5.3  
12:00 228.3 F 9.3  
14:00 236.0 F 12.0  
16:00 239.3 F 11.1  
18:00 236.2 F 9.3  
20:00 239.3 F 9.8  
22:00 249.4 F 15.1  
24:00 261.1 F 21.7  
26:00 271.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.5 F 29.1  
2:00 271.5 F 28.7  
3:00 272.4 F 29.1  
4:00 271.5 F 28.5

Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
5:00 202.2 F 0.0  
10:00 205.9 F 0.0  
15:00 203.8 F 0.0  
20:00 196.6 F 0.0  
25:00 193.3 F 0.0  
30:00 209.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:37  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:04:49  
CYCLE COMPLETE

07/31/2023 04:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_