

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2721

Name: RITTER2  
Sterilizer ID: V2313228  
08/01/2023 03:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
2:00 211.5 F 4.0  
4:00 207.6 F 9.8  
6:00 221.2 F 6.4  
8:00 235.4 F 11.8  
10:00 236.4 F 9.9  
12:00 236.2 F 8.7  
14:00 251.2 F 16.2  
16:00 263.3 F 23.4  
18:00 272.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.8 F 29.8  
2:00 271.9 F 29.1  
3:00 271.6 F 29.3  
4:00 272.7 F 29.8  
  
Min 271.4 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 216.4 F 0.0  
10:00 235.9 F 0.0  
15:00 224.5 F 0.0  
20:00 222.7 F 0.0  
25:00 221.9 F 0.0  
30:00 221.8 F 0.0  
Drying Complete

FILLING: 3:03  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:27  
CYCLE COMPLETE

08/01/2023 04:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_