

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2346

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2023 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.8 F 0.0  
2:00 160.1 F 1.3  
4:00 163.7 F 4.2  
6:00 193.4 F 8.7  
8:00 211.4 F 11.3  
10:00 224.1 F 8.4  
12:00 236.1 F 13.5  
14:00 238.0 F 10.9  
16:00 236.7 F 9.3  
18:00 250.1 F 15.8  
20:00 261.4 F 22.3  
22:00 271.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.9 F 29.7  
2:00 271.6 F 29.0  
3:00 272.0 F 29.6  
4:00 272.4 F 29.5

Min 271.2 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 214.1 F 0.0  
10:00 212.0 F 0.0  
15:00 212.5 F 0.0  
20:00 213.8 F 0.0  
25:00 217.1 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:36  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:01:14  
CYCLE COMPLETE

05/10/2023 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_