
Midmark M11 - v1.0.0
Total Cycles: 2248

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 201.1 F 0.7
4:00 215.0 F 3.9
6:00 230.3 F 9.3
8:00 236.1 F 9.0
10:00 241.5 F 12.1
12:00 240.3 F 10.6
14:00 238.5 F 9.5
16:00 253.5 F 16.8
18:00 265.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.4 F 28.5
2:00 271.9 F 28.4
3:00 272.3 F 28.4
4:00 271.8 F 28.3

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 240.0 F 0.0
10:00 231.7 F 0.0
15:00 228.3 F 0.0
20:00 227.1 F 0.0
25:00 229.3 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

08/25/2022 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____