

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3073

Name: RITTER2  
Sterilizer ID: V2313228  
11/01/2023 04:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.3 F 0.0  
2:00 137.9 F 1.2  
4:00 154.4 F 4.3  
6:00 184.6 F 8.2  
8:00 199.4 F 12.8  
10:00 219.9 F 6.7  
12:00 229.0 F 11.4  
14:00 231.3 F 8.3  
16:00 240.5 F 11.7  
18:00 244.6 F 13.2  
20:00 256.9 F 19.7  
22:00 267.9 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 273.0 F 29.7  
2:00 271.3 F 29.0  
3:00 272.4 F 29.9  
4:00 272.2 F 29.5

Min 271.0 F 29.0  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 228.8 F 0.0  
10:00 201.4 F 0.0  
15:00 198.8 F 0.0  
20:00 197.0 F 0.0  
25:00 195.8 F 0.0  
30:00 196.0 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 23:19  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:02:08  
CYCLE COMPLETE

11/01/2023 05:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_