
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000032
12:13 09/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:33
STERILIZING 3:28
136.0 C 220kPa 3:28
135.8 C 220kPa 3:59
136.3 C 224kPa 4:29
136.3 C 224kPa 4:59
136.2 C 223kPa 5:29
136.4 C 225kPa 5:59
136.4 C 222kPa 6:29
136.4 C 224kPa 6:59
136.4 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 7:52

Digital Signature #
8A02B5896F1FFE13
