
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001132
12:59 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:31
STERILIZING 5:48
136.0 C 220kPa 5:48
136.3 C 224kPa 6:18
136.3 C 224kPa 6:48
136.4 C 224kPa 7:18
136.4 C 224kPa 7:48
136.4 C 225kPa 8:18
136.3 C 224kPa 8:48
136.3 C 223kPa 9:18
136.3 C 223kPa 9:18
VENTING 9:18
AIR DRYING 10:14
CYCLE COMPLETE 21:11

Digital Signature #
0902FDA15C8466D1
