
Midmark M11 - v1.0.5
Total Cycles: 769

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.0 F 0.0
2:00 145.6 F 2.4
4:00 169.8 F 6.1
6:00 194.3 F 10.8
8:00 214.2 F 4.7
10:00 225.5 F 10.5
12:00 230.4 F 7.4
14:00 243.1 F 13.5
16:00 243.2 F 12.4
18:00 255.6 F 18.9
20:00 266.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.9 F 29.6
2:00 272.4 F 29.5
3:00 271.4 F 29.0
4:00 272.1 F 29.9

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 217.7 F 0.0
10:00 219.9 F 0.0
15:00 214.4 F 0.0
20:00 215.5 F 0.0
25:00 213.6 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

02/08/2022 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____