
Midmark M11 - v1.0.0
Total Cycles: 1621

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 133.1 F 0.0
4:00 167.0 F 0.0
6:00 197.4 F 0.0
8:00 212.5 F 0.4
10:00 219.4 F 2.3
12:00 227.5 F 5.1
14:00 235.6 F 8.3
16:00 244.1 F 12.1
18:00 238.4 F 9.5
20:00 239.9 F 9.7
22:00 245.4 F 12.2
24:00 246.1 F 13.1
26:00 259.6 F 20.4
28:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.9 F 28.8
2:00 271.9 F 28.0
3:00 272.4 F 28.7
4:00 272.6 F 28.5

Min 271.5 F 27.8
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 237.4 F 0.0
10:00 224.0 F 0.0
15:00 199.6 F 0.0
20:00 192.2 F 0.0
25:00 191.3 F 0.0
30:00 192.9 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 28:36
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:10
CYCLE COMPLETE

02/10/2022 07:29 PM
ACCEPTED: X
REJECTED:
Operator: U02