

SME12381 — Piedmont Imaging at Piedmont West MRI

Piedmont HealthCare · Siemens MRI · Atlanta, GA · Jul 7 – Aug 6 monitor data (1,468 captures) · analyzed 2026-08-06

COMPRESSOR

ON

running continuously through the window

COLDHEAD

42.300 K

steady at base temperature

PRESSURE NOW

15.30 PSI

within the 14.4–16.4 PSI band · rising

EVENT PEAK

15.32 PSI


Jul 8 00:15Z · 1.08 PSI under threshold

HELIUM


62.3%

+0.00 pts · no loss · no quench

HE PRESSURE — JUL 7 → AUG 6 (daily range, PSI)



HELIUM LEVEL — JUL 7 → AUG 6 (daily reading, %)



Timeline (UTC): Baseline Jul 7 – Aug 6: coldhead at base temperature. No compressor events, alarms, or threshold breaches detected across 1,468 captures. Helium rose to **62.3%** (+0.00 vs baseline). No quench. Pressure now reads 15.30 PSI (Aug 6 13:45Z) — within the 14.4–16.4 PSI band.

CURRENT — AUG 6 (last readings)

Compressor ON. He pressure 15.30 PSI (13:45Z). Helium 62.3%.

RATES

Pressure +0.00 PSI vs baseline (15.30 → 15.30); peak 15.32. Helium +0.00 points.

DATA NOTES

1,468 captures · 1,468 valid pressure · 1,468 valid helium readings. Charts show the daily min-max range with the daily last value. EDU (1,468 captures): room 78.4–78.4 °F · probes 76.2–76.2 / 85.7–85.7 °F · humidity 42–42%.