

SME15822 — Premier Radiology - Smyrna MRI

Althea USA · Philips MRI · Smyrna, TN · Jun 28 – Jul 22 monitor data (1,171 captures) · analyzed 2026-08-06

COMPRESSOR

ON

recovered Jul 22
03:48Z · 2 cycles
from Jul 20 13:48Z

PRESSURE NOW

262 mbar

~3.3× the 80 mbar
line · easing

EVENT PEAK

282 mbar

Jul 21 08:48Z · 202
mbar over threshold

TEMP ALARM

TRIGGERED


27 readings · Jul 20
13:48Z → Jul 21 04:48Z

HELIUM


79.3%

+2.90 pts · no loss ·
no quench

HE PRESSURE — JUN 28 → JUL 22 (daily range, mbar · orange shade = the event window)



HELIUM LEVEL — JUN 28 → JUL 22 (daily reading, %)



Timeline (UTC): Baseline Jun 28 - Jul 20: pressure held 30–37 mbar, helium 76.4–76.5%. **Jul 20 13:48Z: the compressor stopped** and then cycled on/off 2 times through Jul 22 03:48Z (28 readings off). A compressor temperature alarm was triggered Jul 20 13:48Z → Jul 21 04:48Z (27 readings — not continuous, it re-triggered 3 times; amber strip on the pressure chart). Room temperature held 73.6–74.8 °F over the alarm window. He pressure rose from 30 to a **peak of 282 mbar at Jul 21 08:48Z** — 202 mbar over the 80 mbar line (+13 mbar/h avg on the ramp). **The compressor recovered Jul 22 03:48Z** and has held since. Helium rose to **79.3%** (+2.90 vs baseline). No quench. Pressure now reads 262 mbar (Jul 22 08:48Z) — still 182 mbar over the line.

CURRENT — JUL 22 (last readings)

Compressor ON. He pressure 262 mbar (08:48Z). Helium 79.3%.

RATES

Pressure +232 mbar vs baseline (30 → 262); peak 282. Ramp averaged +13 mbar/h. Helium +2.90 points.

DATA NOTES

1,171 captures · 49 valid pressure · 52 valid helium readings. Charts show the daily min-max range with the daily last value. Tech Room Temp 18.7–21.7 °C in window. EDU (1,199 captures): room 72.9–80.2 °F · probes 60.7–70.0 / 65.5–80.3 °F · humidity 40–72%. Host clock ~62 min off vs capture time — times shown use capture time.