# SART - Safety Assessment, Reporting and Tracking Template

Case opening, purpose and awareness

SIMULATED · v1.0

| | | | |
|---|---|---|---|
| **Study / protocol**<br>GLENOS-01 / GS-2026-01 | **Current CIP / version**<br>CIP v4.2 | **Site / code**<br>Riverside Orthopedic Center / 014 | **SART case ID / revision**<br>SART-014-002 / r2 |
| **Participant code (or N/A and why)**<br>014-0002 | **Procedure, visit or exposure (or N/A)**<br>Reverse shoulder arthroplasty, 2 Oct 2026 (ICDL-014-002) | **Site awareness: date / time**<br>2 Oct 2026, 10:42 | **Person who received the first information**<br>Dr R. Teixeira, operating surgeon |
| **PI / authorized site assessor**<br>Dr Alex Morgan | **Sponsor safety system / route**<br>GLENOS EDC, adverse event page | **Sponsor safety case ID (if issued)**<br>GS-AE-014-0007 | **Current local case status**<br>☐ Open ☐ Awaiting sponsor<br>☑ Closed locally |

## A Purpose and use

The SART is the site's record of one safety case, from the moment the site becomes aware of it until the site's part of the case is closed. It shows that the site recognized the information, protected the person, made its own assessment, reported through the sponsor's route within the CIP window, followed the case and linked it to its other records.

Open the SART at awareness, without waiting for a complete classification. Sections B to D are completed in the first days; Section E grows with each new piece of information; Sections F and G close the site's part.

**Boundary.** This is a controlled site coordination record. It does not replace source records, the sponsor safety form, the eCRF, the sponsor safety portal, the device-accountability record, the deviation record, the SCML, the CSST, the CAPA log, a regulatory submission or the sponsor/CEC safety determination. It records the site's part of the case and the references to those records.

| | |
|---|---|
| **Unit of analysis** | One safety case at the site: one adverse event, one observed device deficiency or one piece of new safety information. |
| **Decision supported** | Whether the site recognized, protected, assessed, reported, followed and closed its part of the case in a controlled way. |
| **Inputs** | Source records, participant or user contact, device and accountability records, procedure or visit records (ICDL, PFVR, IQL), sponsor communications, the current CIP and risk information. |
| **Outputs** | The local trail of the case; references to the sponsor system; the reporting data read by the PIR (safety reporting timeliness); the links to the SCML, the deviation record and, when triggered, the CSST, POT, SRR or CAPA. |

**Rule:** if an adverse event is associated with an observed device deficiency, apply both the adverse-event pathway and the device-deficiency pathway (Section C).

## B Awareness, immediate protection and source map

Record what is known at awareness and where it is documented. Do not copy narratives from the clinical record or the sponsor form.

| Item | Status | Record / reference |
|---|---|---|
| **Occurrence, deficiency detection or new information: date / time** | ☑ Known ☐ Estimated ☐ Not yet known | 2 Oct 2026, 10:40, during humeral preparation |
| **Site awareness: date / time, and who brought the information (person and source)** | | 2 Oct 2026, 10:42; Dr R. Teixeira, in theatre; ICDL-014-002 Section G |
| **Initial trigger** | ☑ Adverse event<br>☐ Observed device deficiency<br>☐ AE associated with an observed DD<br>☐ New safety information<br>☐ Other CIP-defined trigger | Intraoperative event |
| **Immediate safety concern for the participant, user or another person** | ☐ Yes ☑ No ☐ Not yet determined | Fracture fixed; participant stable at closure |
| **Immediate action taken: clinical care, device containment or communication (who, when)** | | Fixed with two transosseous sutures; PI called before closure (POT trigger 10), 10:55 |
| **Device and procedure context: system and component; model, lot, serial or UDI; size; technical step (or N/A)** | | GLENOS v2 humeral stem 12 mm, lot GS-H12-2609; no device implicated; humeral preparation step |

# SART - Safety Assessment, Reporting and Tracking Template

Awareness and immediate protection

SIMULATED · v1.0

| Study / site | SART case ID / revision | Participant code or N/A | Site awareness: date / time |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-002 / r2 | 014-0002 | 2 Oct 2026, 10:42 |

## B Awareness, immediate protection and source map (continued)

**Factual summary (what happened, as known now)**

During humeral preparation in case 2, a non-displaced greater tuberosity fracture was seen. It was fixed with two transosseous sutures and the planned implant was placed. The shoulder was stable at closure, and the immediate postoperative radiograph shows the implant and the fixation in position.

**Source map**

| Source record | ID / location | What it holds for this case |
|---|---|---|
| Operative note | EHR, 2 Oct 2026 | Fracture, fixation and final configuration |
| ICDL | ICDL-014-002, Sections F and G | Time of the event and of the call to the PI |
| Immediate postoperative image | IQL-014-002-PO | Implant and fixation position |
| 6-week visit record | PFVR-014-002-V1 | Healing of the tuberosity |
| eCRF | GLENOS EDC, AE-07 | Report to the sponsor |
| | | |

**Missing information**

| Information still missing | Owner (Role ID) | Next step / due |
|---|---|---|
| 6-week radiograph result | P-04 | At the 6-week visit, 13 Nov 2026 |
| | | |
| | | |
| | | |
| | | |

# SART - Safety Assessment, Reporting and Tracking Template

Current PI / site assessment

SIMULATED · v1.0

| Study / site | SART case ID / revision | Participant code or N/A | Assessment date |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-002 / r2 | 014-0002 | 2 Oct 2026 (PI) |

## C Current PI / site assessment

This is the site's assessment and may change. The sponsor or CEC determination is recorded apart, in the last table. Rationale and source go in the right-hand column.

| C1 · Adverse-event pathway | | |
|---|---|---|
| **Adverse event identified?** | ☑ Yes ☐ No ☐ Pending | |
| **Seriousness (mark all that apply)** | ☐ Death<br>☐ Life-threatening illness or injury<br>☐ Permanent impairment of a body structure or function<br>☐ In-patient or prolonged hospitalization<br>☐ Intervention to prevent the above<br>☐ Fetal distress, fetal death or congenital abnormality<br>☑ Does not meet seriousness criteria<br>☐ Pending | |
| **Clinical severity (intensity, on the CIP scale): kept apart from seriousness** | ☑ Mild ☐ Moderate ☐ Severe ☐ Pending | CIP v4.2 intensity scale |
| **Related to the investigational device?** | ☐ Yes ☑ No ☐ Pending | Bone event during preparation; implant not involved |
| **Related to the comparator?** | ☐ Yes ☐ No ☐ Pending ☑ N/A | Single-arm study |
| **Related to a CIP-required procedure outside routine practice?** | ☐ Yes ☑ No ☐ Pending ☐ N/A | Implantation is the study procedure, not an added one |
| **If a SADE: identified in the current risk assessment?** | ☐ ASADE ☐ USADE ☐ Pending ☑ N/A | Not a SADE |
| **Resulting local category** | ☑ Non-serious AE ☐ SAE ☐ ADE ☐ SADE ☐ ASADE ☐ USADE ☐ Pending | |

| C2 · Device-deficiency pathway | | |
|---|---|---|
| **Observed device deficiency?** | ☐ Yes ☑ No ☐ Pending | No deficiency observed |
| **Type (mark all that apply)** | ☐ Identity ☐ Quality ☐ Durability ☐ Reliability ☐ Usability ☐ Safety ☐ Performance ☐ Malfunction ☐ Use error ☐ Information or label | |
| **Associated with an adverse event? (Yes: apply both pathways)** | ☐ Yes ☐ No ☐ Pending | |
| **Could it have led to a SADE?** | ☐ Yes ☐ No ☐ Pending | |
| **Device containment, return or analysis (reference)** | ☐ Done ☐ Pending ☑ N/A | |

| C3 · Context and urgency | | |
|---|---|---|
| **Contributing context, as now assessed (for the SCML; mark all that apply)** | ☐ Device ☑ Procedure ☑ Operator or technique ☐ Patient ☐ Undetermined | Early case in the surgeon's sequence (SLCT case 2) |
| **Serious health threat (imminent risk needing rapid action to protect others)?** | ☐ Yes ☑ No ☐ Pending | |

| C4 · Who has assessed the case | | |
|---|---|---|
| **Current PI / site assessment: status, assessor, date** | ☐ Pending ☐ Preliminary ☑ Final (PI) | A. Morgan, 2 Oct 2026 |
| **Sponsor / CEC determination: date and reference** | ☐ Not yet received ☑ Received | 20 Oct 2026, MM-Q-114 closed |
| **Agreement between the two** | ☑ Agree ☐ Disagree ☐ Not yet known | Non-serious, not device-related |
| **Open classification question and owner** | None | |

# SART - Safety Assessment, Reporting and Tracking Template

Reporting to the sponsor

SIMULATED · v1.0

| Study / site | SART case ID / revision | Participant code or N/A | Sponsor safety case ID |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-002 / r2 | 014-0002 | GS-AE-014-0007 |

## D Reporting to the sponsor

Record the proof of reporting, not the content of the sponsor form. Every time limit comes from the CIP or the sponsor procedure. The PIR reads its safety reporting timeliness counts from this section.

| Reporting step | Status | Record / reference |
|---|---|---|
| **Reportable under the current CIP or sponsor procedure? (rule reference)** | ☑ Yes ☐ No ☐ Pending | CIP v4.2 section 9.3: every AE entered in the eCRF within 5 working days |
| **Route used** | ☐ Sponsor portal ☑ eCRF ☐ SAE form ☐ DD form ☐ Emergency route ☐ Other | GLENOS EDC adverse event page |
| **Site awareness: date / time (from Section B)** | | 2 Oct 2026, 10:42 |
| **CIP reporting window, and due date / time** | | 5 working days; due 9 Oct 2026 |
| **Initial notification sent: date / time, by whom, system reference** | | 3 Oct 2026, 16:10; M. Silva (P-06); EDC entry AE-07 |
| **Receipt confirmed / sponsor safety case ID** | ☑ Yes ☐ Pending | GS-AE-014-0007 |
| **Initial notification on time?** | ☑ Yes ☐ No ☐ Pending ☐ N/A | Sent 1 working day after awareness |
| **First report complete?** | ☑ Complete ☐ Incomplete ☐ Pending ☐ N/A | No query on the first report |
| **If late or incomplete: where the delay arose** | ☐ Recognition ☐ Internal escalation ☐ PI assessment ☐ Access to documentation ☐ Transmission ☐ Other | N/A: on time and complete |
| **Correction made, and deviation record ID if the notification was late** | | N/A |
| **Ethics committee or authority communication, only if the CIP assigns it to the site** | ☐ Sent ☑ Not required ☐ Pending | Not assigned to the site by the CIP |

# SART - Safety Assessment, Reporting and Tracking Template

Follow-up and new information

SIMULATED · v1.0

| Study / site | SART case ID / revision | Participant code or N/A | Sheet number / total sheets |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-002 / r2 | 014-0002 | 1 / 1 |

## E Follow-up and new information

One line for each new piece of information or sponsor request. Add lines; never overwrite an earlier one.

| No. | Date / time received | New clinical, technical or device information, and its source | Effect on the site assessment | Sponsor request or communication (reference) | Follow-up due | Sent / receipt reference | On time? | Owner (Role ID) and next step |
|---|---|---|---|---|---|---|---|---|
| 1 | 6 Oct 2026, 09:00 | Read at the scheduled oversight review with SLCT case 2; no pattern (POT-014-R-2026-10) | None | N/A | N/A | N/A | ☐ Yes ☐ No ☑ N/A | P-01: SLCT review at case 5 |
| 2 | 14 Oct 2026, 11:20 | Sponsor medical monitor asks for the fixation details (MM-Q-114) | None | MM-Q-114 | 21 Oct 2026 | Answered 15 Oct; EDC receipt | ☑ Yes ☐ No ☐ N/A | P-06: await the determination |
| 3 | 20 Oct 2026, 15:05 | Sponsor confirms: non-serious, not device-related; MM-Q-114 closed | Confirms the PI assessment | MM-Q-114 closure | N/A | N/A | ☐ Yes ☐ No ☑ N/A | P-06: record in C4 |
| 4 | 13 Nov 2026, 10:30 | 6-week visit: tuberosity healed on radiograph (PFVR-014-002-V1; IQL-014-002-W6) | Outcome: recovered | N/A | 20 Nov 2026 | Sent 14 Nov; EDC AE-07 v2 | ☑ Yes ☐ No ☐ N/A | P-06: close locally |
| 5 | | | | | | | ☐ Yes ☐ No ☐ N/A | |
| 6 | | | | | | | ☐ Yes ☐ No ☐ N/A | |
| 7 | | | | | | | ☐ Yes ☐ No ☐ N/A | |
| 8 | | | | | | | ☐ Yes ☐ No ☐ N/A | |

If more lines are needed, reprint this sheet, number it and continue the numbering.

# SART - Safety Assessment, Reporting and Tracking Template

Links and closure

SIMULATED · v1.0

| Study / site | SART case ID / revision | Participant code or N/A | Local closure date |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-002 / r2 | 014-0002 | 16 Nov 2026 |

## F Records this case is sent to

Each line is a record that receives this case from the SART. Mark whether the link was made and write the receiving record's ID.

| Record | When it applies | ID / date | Status |
|---|---|---|---|
| **SCML** | Every adverse event or device deficiency (DD) at the site · SCML event ID | SCML-014-001, 2 Oct 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **Deviation record** | The initial notification was late; the sponsor's criteria decide its importance · Deviation ID | | ☐ Linked ☑ Not required ☐ Pending |
| **CSST, early review** | Serious health threat, cluster, new category, or concentration in one component or surgeon · Review reference | No trigger | ☐ Linked ☑ Not required ☐ Pending |
| **POT** | The case, a delay or a pending decision needs the PI's oversight · POT review ID | POT-014-R-2026-10, 6 Oct 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **SRR or CAPA** | A recurring failure or a weak local control; not for a single case · Record ID | Single case | ☐ Linked ☑ Not required ☐ Pending |
| **PIR** | Every reportable case enters the safety reporting timeliness count of its review period · PIR ID / period | PIR-014-2026-10, October 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **Sponsor device-deficiency route** | The device deficiency needs return, analysis or action by the sponsor · Sponsor DD case ID | No deficiency | ☐ Linked ☑ Not required ☐ Pending |

## G Closing the site's part of the case

| Before closing, confirm that… | Result | Reference / limitation |
|---|---|---|
| Immediate protection and care documented (or N/A, with the reason) | ☑ Yes ☐ No ☐ N/A | Clinical record, 2 Oct |
| Source references enough to reconstruct the case | ☑ Yes ☐ No ☐ N/A | Section B source map |
| Current PI / site assessment documented | ☑ Yes ☐ No ☐ N/A | C4 |
| Sponsor submission and receipt documented | ☑ Yes ☐ No ☐ N/A | D: AE-07 / GS-AE-014-0007 |
| Timeliness and completeness known, or the limitation recorded | ☑ Yes ☐ No ☐ N/A | D: on time, complete |
| Follow-up completed, transferred or explicitly open | ☑ Yes ☐ No ☐ N/A | E line 4 |
| Sponsor / CEC determination recorded, or recorded as pending | ☑ Yes ☐ No ☐ N/A | C4: 20 Oct 2026 |
| Links in Section F completed | ☑ Yes ☐ No ☐ N/A | F |

**Local closure status**

☑ Closed locally ☐ Open follow-up ☐ Transferred ☐ Not yet eligible

**Current outcome and open items**

Recovered; tuberosity healed at 6 weeks. No open items.

Local closure does not close the sponsor, CEC, ethics committee or regulatory process.

### Case reporter / initial site assessor

Name / role: Dr R. Teixeira / Sub-investigator

Signature / date:

R. Teixeira 2 Oct 2026

### PI / accountable investigator

Name / role: Dr Alex Morgan / PI

Signature / date:

A. Morgan 16 Nov 2026

### Safety coordinator (if applicable)

Name / role: Maria Silva / Study Coordinator

Signature / date:

M. Silva 16 Nov 2026

**Controlled location of the SART and linked system references:** Investigator site file, tab 6 (Safety) / SART-014-002; GLENOS EDC AE-07; SCML-014