IQL — IMAGING QUALITY LOG

v0.5 - SIMULATED

STUDY / PROTOCOL / CIP VERSION	SITE / CODE	IQL RECORD ID / REVISION	REVIEW PERIOD / CUT-OFF DATE
NB-Shoulder-01 / CIP v3.2	Northbridge Orthopaedic Site / NB-014	IQL-014 / r1	01–30 Sep 2026; cut-off 30 Sep

APPLICABLE IMAGING CHARTER / MANUAL ID AND VERSION	INCLUDED MODALITY / TIME-POINT SCOPE
IM-NB-01 v1.2 / sponsor Imaging Manual	Shoulder radiographs: screening and Day 30 follow-up
PREPARED BY / DATE	**CONTROLLED RECORD LOCATION**
Elena Martin / Imaging Lead; 30 Sep 2026	Site file / Imaging / IQL-014; PACS source records

A Purpose and Use

Record each required image set at acquisition or first availability, using the applicable sponsor Imaging Charter or imaging manual. Document acquisition timing, required-set completeness, technical quality, the local use decision and retrievable source evidence.

Link to PSEAR (screening and eligibility) or PFVR (follow-up) as applicable. Clinical interpretation, eligibility decisions and sponsor or central endpoint review remain in their authoritative records.

Element	Scope / Decision	Record / Link
Unit of analysis	One required image set for one participant and one protocol time point.	IQL Record ID + image-set ID; participant, time point and linked PSEAR/PFVR ID.
Decision supported	Whether the set is complete, technically adequate and locally usable for its defined purpose; identify any limitation, repeat or second review needed.	Record the four classifications in the image-set log. Pending review is not a positive use decision.
Inputs	Applicable CIP, sponsor imaging requirements, acquisition metadata, required views or sequences, source images and authorised review evidence.	Charter/manual ID and version above; set-specific requirement and archive/transfer reference in the image-set log.
Outputs	Traceable image-set status for PSEAR/PFVR and a linked response for unresolved or recurring quality issues.	Image-set resolution record; SART, deviation, IMC, SRR or CAPA ID when applicable. FUCR receives visit evidence from PFVR.

B Image-Set Classification Guide

Illustrations only, not acceptance thresholds. Apply the approved Imaging Charter/manual. Assess each dimension separately; these examples are a reminder, not fields to complete.

Dimension	What to distinguish	Example / recording reminder
Timing	Assess acquisition against its applicable window; retain the actual date.	Window ends 24 Sep; acquired 26 Sep: Outside. Later recovery does not make the original acquisition timely.
Set completeness	Compare required views or sequences with what was actually acquired.	Two of three required views: Partial. Technically clear images do not establish a complete set.
Technical quality	Assess positioning, coverage and other sponsor-defined technical criteria.	All views present, but positioning limits assessment: Limited. No images available: Not assessed.
Local use decision	Record whether the set can support its stated local purpose, including any restriction.	A necessary second review is outstanding: Pending. Limited use needs a documented permitted purpose; it does not establish central acceptance.

IQL — IMAGING QUALITY LOG

v0.5 - SIMULATED

STUDY / SITE CODE	IQL RECORD ID / REVISION	IMAGING CHARTER / MANUAL VERSION	REVIEW PERIOD / SHEET NO.
NB-Shoulder-01 / NB-014	IQL-014 / r1	IM-NB-01 v1.2	Sep 2026 / sheet 1 of 1

C Image-Set Quality Log Against Sponsor Imaging Requirements

One row per image set. Link PSEAR/PFVR when applicable. Use Resolution / Route ID only for an exception, repeat, second review or other response; otherwise enter N/A. Do not rely on a set for a use that remains pending. Local acceptance does not establish central acceptance.

Set ID	Participant / Time Point	PSEAR / PFVR ID	Modality / Required Set / Requirement Ref.	Planned Window / Actual Acquisition Date	Timing	Required Set Complete?	Technical Quality	Local Use Decision	Archive / Transfer / Receipt Ref.	Reviewer / Date	Resolution / Route ID
01	NB014-0031 / Screening	PSEAR-NB014-0031	XR: AP/Y/axillary; IM-NB-01 4.2	Window 10-14 Sep; acquired 12 Sep	☒ In window ☐ Outside ☐ Pending ☐ N/A	☒ Complete ☐ Partial ☐ Missing ☐ N/A	☒ Adequate ☐ Limited ☐ Not conforming ☐ Not assessed	☒ Usable ☐ Limited use ☐ Pending ☐ Not usable	IMG-NB014-0031-SCR; TX-01	E. Martin / 12 Sep	N/A
02	NB014-0032 / Day 30	PFVR-NB014-0032-D30	XR: AP/Y/axillary; IM-NB-01 4.2	Window 10-24 Sep; acquired 26 Sep	☐ In window ☒ Outside ☐ Pending ☐ N/A	☒ Complete ☐ Partial ☐ Missing ☐ N/A	☒ Adequate ☐ Limited ☐ Not conforming ☐ Not assessed	☐ Usable ☒ Limited use ☐ Pending ☐ Not usable	IMG-NB014-0032-D30; TX-02	E. Martin / 26 Sep	R01; DEV-013
03	NB014-0033 / Day 30	PFVR-NB014-0033-D30	XR: 3 views required; axillary missing	Window 25 Sep-02 Oct; acquired 29 Sep	☒ In window ☐ Outside ☐ Pending ☐ N/A	☐ Complete ☒ Partial ☐ Missing ☐ N/A	☒ Adequate ☐ Limited ☐ Not conforming ☐ Not assessed	☐ Usable ☐ Limited use ☒ Pending ☐ Not usable	IMG-NB014-0033-D30; TX-03	E. Martin / 29 Sep	R02
04	NB014-0034 / Day 30	PFVR-NB014-0034-D30	XR: 3 views; positioning limitation	Window 24-30 Sep; acquired 28 Sep	☒ In window ☐ Outside ☐ Pending ☐ N/A	☒ Complete ☐ Partial ☐ Missing ☐ N/A	☐ Adequate ☒ Limited ☐ Not conforming ☐ Not assessed	☐ Usable ☒ Limited use ☐ Pending ☐ Not usable	IMG-NB014-0034-D30; TX-04	E. Martin / 29 Sep	R03
05	NB014-0035 / Day 30	PFVR-NB014-0035-D30	XR: 3 views; motion obscures landmarks	Window 25 Sep-02 Oct; acquired 29 Sep	☒ In window ☐ Outside ☐ Pending ☐ N/A	☒ Complete ☐ Partial ☐ Missing ☐ N/A	☐ Adequate ☐ Limited ☒ Not conforming ☐ Not assessed	☐ Usable ☐ Limited use ☐ Pending ☒ Not usable	IMG-NB014-0035-D30; TX-05	E. Martin / 29 Sep	R02

Continue image-set IDs on additional log sheets. Use IQL Record ID + set ID for unique links. Retain the original entry and cross-reference any repeat acquisition.

IQL — IMAGING QUALITY LOG

v0.5 - SIMULATED

STUDY / SITE CODE	IQL RECORD ID / REVISION	IMAGING CHARTER / MANUAL VERSION	REVIEW PERIOD / SHEET NO.
NB-Shoulder-01 / NB-014	IQL-014 / r1	IM-NB-01 v1.2	Sep 2026 / sheet 1 of 1

D Periodic Imaging-Quality Summary and Handoff

Summarise the reviewed image sets. Record the distinct-set count and relevant IDs; categories may overlap. Refer recurring or unresolved issues to the owning record without duplicating FUCR indicators or a CAPA plan.

IMAGE-SET IDS / SCOPE REVIEWED	DISTINCT IMAGE SETS REVIEWED (N)
IQL-014 / sets 01-05; screening + Day 30	5 distinct required image sets

Quality Aspect	Count / N and Image-set IDs	Finding / No Further Action or Linked Route
Set completeness — partial or missing	1/5 - set 03	Missing axillary view. R02 repeat request; PFVR-0033 remains pending for this component.
Technical quality — limited or non-conforming	2/5 - sets 04, 05	Set 04: limited positioning. Set 05: non-conforming motion. Different issues; no common-cause conclusion.
Local use — pending or not usable	2/5 - sets 03, 05	Set 03 Pending; set 05 Not usable. R02 open, due 01 Oct. Sets 02/04 have purpose-specific limited use.
Repeat acquisition / second review required during period	3/5 - sets 03, 04, 05	Repeat for 03/05 remains open. Second review of 04 completed 29 Sep (R03). Counts are not additive.

E Image-Set Resolution and Linked Route

Record immediate resolution and evidence of the resulting use decision. Cite a CAPA, risk, deviation or safety record when that route is needed. Keep this response open until the local resolution or required handoff is documented.

Resolution ID / Set ID / Issue / Use Impact	Immediate Response / Repeat / Second Review	Resolution Evidence / Updated Local Use / Date	Linked Participant / Governed Record	Owner / Due Date	Status
R01 / set 02: complete, adequate, but 2 days late.	Radiographs recovered 26 Sep; PI reviewed the clinical report. Original late status retained.	26 Sep: local clinical review permitted; endpoint timing criterion not met. IMG-0032-D30 / DEV-013.	PFVR-NB014-0032-D30; DEV-NB014-2026-013	E. Martin / 26 Sep	☐ Open ☒ Resolved ☐ Handed off
R02 / sets 03, 05: missing view / motion; required use unresolved.	Repeat requested within the original windows. Keep the two participant records and repeat sources distinct.	30 Sep: repeat images not yet available. 03 Pending; 05 Not usable. No retrospective date change.	PFVR-NB014-0033-D30; PFVR-NB014-0035-D30	E. Martin / 01 Oct	☒ Open ☐ Resolved ☐ Handed off
R03 / set 04: complete but limited positioning.	Second local review by Dr Sofia Chen on 29 Sep; document the permitted scope and restriction.	Clinical comparison permitted; quantitative endpoint assessment remains restricted. REV-0034-D30; no central acceptance inferred.	PFVR-NB014-0034-D30; REV-0034-D30	Dr Sofia Chen / 29 Sep	☐ Open ☒ Resolved ☐ Handed off

IQL — IMAGING QUALITY LOG

v0.5 - SIMULATED

STUDY / SITE CODE	IQL RECORD ID / REVISION	IMAGING CHARTER / MANUAL VERSION	REVIEW PERIOD / SHEET NO.
NB-Shoulder-01 / NB-014	IQL-014 / r1	IM-NB-01 v1.2	Sep 2026 / sheet 1 of 1

F Current Period Imaging-Quality Conclusion

Select the status reflecting the highest unresolved issue in the reviewed scope. This conclusion concerns local imaging quality and outstanding routes, not participant eligibility or central endpoint adjudication.

Conclusion	Meaning / Response	Select One	Conclusion Date / Next Review
Controlled / no open issue	Reviewed entries support their documented local use; no unresolved quality issue or handoff is known.	☐ Select	
Controlled with limitation	A documented limitation permits the defined local use, with any required conditions or follow-up identified.	☐ Select	
Resolution / handoff pending	A repeat, second review, traceability correction or linked quality response remains open.	☐ Select	
Use restricted / escalated	At least one set remains restricted for its intended use or requires PI, safety, sponsor or other governed review.	☒ Select	30 Sep 2026 / next 01 Oct

Scope / Affected Image-set IDs / Basis and Open Routes

30 Sep 2026 - EM: R02 remains open for sets 03 and 05; required follow-up imaging use is unresolved. Set 02 is adequate for the documented clinical review but remains outside its endpoint window. Set 04 has a documented clinical-use limitation after second review. No central acceptance or eligibility decision is made here. Update each PFVR from its own image-set status. Review repeat acquisition evidence by 01 Oct; no common-cause or CAPA finding is asserted from these five sets.

G Accountable Local Review and Record Location

The local imaging reviewer signs this review. Obtain PI review when required by the CIP or sponsor instructions, or when an unresolved restriction, repeated failure, safety route or impact on an eligibility/endpoint decision requires PI attention.

Local Imaging Reviewer / Radiology Liaison

Name / Role	Elena Martin / Imaging Lead
Signature	EM - simulated
Date	30 Sep 2026

PI Review / Acknowledgment — When Required

Name / Role	Dr Sofia Chen / PI
Signature	SC - simulated
Date	30 Sep 2026

PI Review Applicability / Trigger or Governing Reference

☒ Required ☐ Not required

Required: unresolved use restrictions (03/05), late CtQ imaging (02), and the local-use limitation in 04. PI acknowledgment recorded; it does not close R02.

NEXT REVIEW DATE / TRIGGER	CONTROLLED LOCATION / IQL REVISION
01 Oct 2026 / review R02 repeats	Site file / Imaging / IQL-014 r1