# SART - Safety Assessment, Reporting and Tracking Template

Case opening, purpose and awareness

SIMULATED · v1.1

| Study / protocol | Current CIP / version | Site / code | SART case ID / revision |
|---|---|---|---|
| GLENOS-01 / GS-2026-01 | CIP v4.2 | Riverside Orthopedic Center / 014 | SART-014-005 / r4 |
| **Participant code (or N/A and why)**<br>014-0005 | **Procedure, visit or exposure (or N/A)**<br>Reverse shoulder arthroplasty, 9 Oct 2026 (ICDL-014-005); revision 28 Oct 2026 | **Site awareness: date / time**<br>28 Oct 2026, 08:05 | **Person who received the first information**<br>Maria Silva (P-06), from the hospital admission list |
| **PI / authorized site assessor**<br>Dr Alex Morgan (on leave 27 Oct to 2 Nov); backup Dr R. Teixeira | **Sponsor safety system / route**<br>GLENOS EDC: SAE form and DD form | **Sponsor safety case ID (if issued)**<br>GS-SAE-014-0003 | **Current local case status**<br>☑ Open ☐ Awaiting sponsor ☐ Closed locally |

## A Purpose and use

The SART is the site's record of one safety case, from the moment the site becomes aware of it until the site's part of the case is closed. It shows that the site recognized the information, protected the person, made its own assessment, reported through the sponsor's route within the CIP window, followed the case and linked it to its other records.

Open the SART at awareness, without waiting for a complete classification. Sections B to D are completed in the first days; Section E grows with each new piece of information; Sections F and G close the site's part.

**Boundary.** This is a controlled site coordination record. It does not replace source records, the sponsor safety form, the eCRF, the sponsor safety portal, the device-accountability record, the deviation record, the SCML, the CSST, the CAPA log, a regulatory submission or the sponsor/CEC safety determination. It records the site's part of the case and the references to those records.

| | |
|---|---|
| **Unit of analysis** | One safety case at the site: one adverse event, one observed device deficiency or one piece of new safety information. |
| **Decision supported** | Whether the site recognized, protected, assessed, reported, followed and closed its part of the case in a controlled way. |
| **Inputs** | Source records, participant or user contact, device and accountability records, procedure or visit records (ICDL, PFVR, IQL), sponsor communications, the current CIP and risk information. |
| **Outputs** | The local trail of the case; references to the sponsor system; the reporting data read by the PIR (safety reporting timeliness); the links to the SCML, the deviation record and, when triggered, the CSST, POT, SRR or CAPA. |

**Rule:** if an adverse event is associated with an observed device deficiency, apply both the adverse-event pathway and the device-deficiency pathway (Section C).

## B Awareness, immediate protection and source map

Record what is known at awareness and where it is documented. Do not copy narratives from the clinical record or the sponsor form.

| Item | Status | Record / reference |
|---|---|---|
| **Occurrence, deficiency detection or new information: date / time** | ☑ Known ☐ Estimated | 27 Oct 2026, 21:30, at home, while dressing (day 18) |
| **Site awareness: date / time, and who brought the information (person and source)** | 28 Oct 2026, 08:05; Maria Silva (P-06), from the hospital admission list; emergency department note | |
| **Initial trigger** | ☑ Adverse event<br>☐ Observed device deficiency<br>☐ AE associated with an observed DD<br>☐ New safety information<br>☐ Other CIP-defined trigger | Dislocation. The device deficiency was found at revision (Section E, line 1) |
| **Immediate safety concern for the participant, user or another person** | ☑ Yes ☐ No ☐ Not yet determined | Dislocated shoulder; participant admitted |
| **Immediate action taken: clinical care, device containment or communication (who, when)** | Closed reduction failed in the emergency department; admitted. Backup PI Dr R. Teixeira informed 28 Oct, 08:40; open revision 28 Oct, 14:00 | |
| **Device and procedure context: system and component; model, lot, serial or UDI; size; technical step (or N/A)** | GLENOS v2; humeral liner 36 mm +3, lot GS-L36-2607 (read from the explant label: missing in ICDL-014-005); assembly of the liner on the stem | |

| Study / site | SART case ID / revision | Participant code or N/A | Site awareness: date / time |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-005 / r4 | 014-0005 | 28 Oct 2026, 08:05 |

## B Awareness, immediate protection and source map (continued)

**Factual summary (what happened, as known now)**

Eighteen days after a reverse shoulder arthroplasty, the participant felt the shoulder give way while dressing and came to the emergency department, where closed reduction failed. At open revision on 28 Oct the humeral liner was found separated from the stem, with the locking ring not engaged. The liner was replaced, the shoulder was stable at closure, and the removed liner was kept for return to the sponsor.

**Source map**

| Source record | ID / location | What it holds for this case |
|---|---|---|
| Emergency department note and admission | EHR, 28 Oct 2026 | Presentation, failed reduction, admission |
| Revision operative note | EHR / ICDL-014-005-R | Liner separated; locking ring not engaged; liner replaced |
| Explant and return record | RMA-0045, 30 Oct 2026 | Liner kept and returned to the sponsor |
| Index procedure log | ICDL-014-005, Section E | Liner lot not recorded at the index procedure |
| Sponsor SAE and DD forms | GS-SAE-014-0003; GS-DD-014-0012 | Formal report and follow-up |
| Device analysis report | DA-0045, 19 Nov 2026 | Locking ring out of tolerance in lot GS-L36-2607 |

**Missing information**

| Information still missing | Owner (Role ID) | Next step / due |
|---|---|---|
| Sponsor lot investigation outcome | P-01 | Sponsor; ask at the next contact |
| 3-month visit after revision | P-06 | Visit window 20 Jan to 3 Feb 2027 |
| | | |
| | | |
| | | |

# SART - Safety Assessment, Reporting and Tracking Template

Current PI / site assessment

SIMULATED · v1.1

| Study / site | SART case ID / revision | Participant code or N/A | Assessment date |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-005 / r4 | 014-0005 | 28 Oct 2026; updated 13 Nov |

## C Current PI / site assessment

This is the site's assessment and may change. The sponsor or CEC determination is recorded apart, in the last table. Rationale and source go in the right-hand column.

| C1 · Adverse-event pathway | | |
|---|---|---|
| **Adverse event identified?** | ☑ Yes ☐ No ☐ Pending | Emergency department note, 28 Oct |
| **Seriousness (mark all that apply)** | ☐ Death<br>☐ Life-threatening illness or injury<br>☐ Permanent impairment of a body structure or function<br>☑ In-patient or prolonged hospitalization<br>☐ Intervention to prevent the above<br>☐ Fetal distress, fetal death or congenital abnormality<br>☐ Does not meet seriousness criteria<br>☐ Pending | Admitted for revision surgery |
| **Clinical severity (intensity, on the CIP scale): kept apart from seriousness** | ☐ Mild ☐ Moderate ☑ Severe ☐ Pending | CIP v4.2 intensity scale |
| **Complication grade (Clavien-Dindo), by the treatment the event needed** | ☐ I ☐ II ☐ IIIa ☑ IIIb ☐ IVa ☐ IVb ☐ V ☐ N/A<br>☐ Pending | Open revision under general anesthesia, 28 Oct |
| **Related to the investigational device?** | ☑ Yes ☐ No ☐ Pending | Liner separation caused the dislocation |
| **Related to the comparator?** | ☐ Yes ☐ No ☐ Pending ☑ N/A | Single-arm study |
| **Related to a CIP-required procedure outside routine practice?** | ☐ Yes ☑ No ☐ Pending ☐ N/A | Event linked to the device, not to an added procedure |
| **If a SADE: identified in the current risk assessment?** | ☐ ASADE ☐ USADE ☑ Pending ☐ N/A | Dislocation is in the risk assessment; liner separation is not |
| **Resulting local category** | ☐ Non-serious AE ☐ SAE ☐ ADE ☑ SADE ☐ ASADE<br>☐ USADE ☐ Pending | Anticipated status for the sponsor |

| C2 · Device-deficiency pathway | | |
|---|---|---|
| **Observed device deficiency?** | ☑ Yes ☐ No ☐ Pending | Found at revision |
| **Type (mark all that apply)** | ☐ Identity ☐ Quality ☐ Durability ☐ Reliability ☐ Usability<br>☐ Safety ☑ Performance ☑ Malfunction ☐ Use error<br>☐ Information or label | Locking ring did not hold the liner |
| **Associated with an adverse event? (Yes: apply both pathways)** | ☑ Yes ☐ No ☐ Pending | Both pathways apply |
| **Could it have led to a SADE?** | ☑ Yes ☐ No ☐ Pending | It did lead to one |
| **Device containment, return or analysis (reference)** | ☑ Done ☐ Pending ☐ N/A | Liner returned, RMA-0045 |

| C3 · Context and urgency | | |
|---|---|---|
| **Contributing context, as now assessed (for the SCML; mark all that apply)** | ☑ Device ☐ Procedure ☑ Operator or technique ☐ Patient<br>☐ Undetermined | Liner seating not documented at the index procedure |
| **Serious health threat (imminent risk needing rapid action to protect others)?** | ☐ Yes ☐ No ☑ Pending | Sponsor checking lot GS-L36-2607 at other sites |

| C4 · Who has assessed the case | | |
|---|---|---|
| **Current PI / site assessment: status, assessor, date** | ☐ Pending ☐ Preliminary ☑ Final (PI) | A. Morgan, 3 Nov 2026 |
| **Sponsor / CEC determination: date and reference** | ☐ Not yet received ☑ Received | 13 Nov 2026, SAF-GS-0031: USADE, device-related |
| **Agreement between the two** | ☑ Agree ☐ Disagree ☐ Not yet known | PI had left the anticipated status to the sponsor |
| **Open classification question and owner** | Operator contribution to liner seating; P-01, after the device analysis | |

# SART - Safety Assessment, Reporting and Tracking Template

Reporting to the sponsor

SIMULATED · v1.1

| Study / site | SART case ID / revision | Participant code or N/A | Sponsor safety case ID |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-005 / r4 | 014-0005 | GS-SAE-014-0003 |

## D Reporting to the sponsor

Record the proof of reporting, not the content of the sponsor form. Every time limit comes from the CIP or the sponsor procedure. The PIR reads its safety reporting timeliness counts from this section.

| Reporting step | Status | Record / reference |
|---|---|---|
| **Reportable under the current CIP or sponsor procedure? (rule reference)** | ☑ Yes ☐ No ☐ Pending | CIP v4.2 section 9.4: SAEs and DDs that could have led to a SADE within 72 hours of awareness |
| **Route used** | ☐ Sponsor portal ☐ eCRF ☑ SAE form ☑ DD form ☐ Emergency route ☐ Other | GLENOS EDC |
| **Site awareness: date / time (from Section B)** | 28 Oct 2026, 08:05 | |
| **CIP reporting window, and due date / time** | 72 hours; due 31 Oct 2026, 08:05 | |
| **Initial notification sent: date / time, by whom, system reference** | 1 Nov 2026, 11:30; M. Silva (P-06); GS-SAE-014-0003 and GS-DD-014-0012 | |
| **Receipt confirmed / sponsor safety case ID** | ☑ Yes ☐ Pending | GS-SAE-014-0003 |
| **Initial notification on time?** | ☐ Yes ☑ No ☐ Pending ☐ N/A | Sent 27 hours late |
| **First report complete?** | ☐ Complete ☑ Incomplete ☐ Pending ☐ N/A | Liner lot missing; sent 2 Nov after the explant label was read |
| **If late or incomplete: where the delay arose** | ☐ Recognition ☑ Internal escalation ☐ PI assessment ☐ Access to documentation ☐ Transmission ☐ Other | Backup PI informed but not told the case was reportable; PI on leave |
| **Correction made, and deviation record ID if the notification was late** | Backup notification rule added to POT Section C; deviation PD-014-12; CAPA-014-2026-004 | |
| **Ethics committee or authority communication, only if the CIP assigns it to the site** | ☐ Sent ☑ Not required ☐ Pending | Not assigned to the site by the CIP |

# SART - Safety Assessment, Reporting and Tracking Template

Follow-up and new information

SIMULATED · v1.1

| Study / site | SART case ID / revision | Participant code or N/A | Sheet number / total sheets |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-005 / r4 | 014-0005 | 1 / 1 |

## E Follow-up and new information

One line for each new piece of information or sponsor request. Add lines; never overwrite an earlier one.

| No. | Date / time received | New clinical, technical or device information, and its source | Effect on the site assessment | Sponsor request or communication (reference) | Follow-up due | Sent / receipt reference | On time? | Owner (Role ID) and next step |
|---|---|---|---|---|---|---|---|---|
| 1 | 29 Oct 2026, 09:10 | Revision findings: liner separated, locking ring not engaged; liner kept (ICDL-014-005-R) | DD pathway opened; device relation confirmed | N/A | N/A | Added to the first report | ☐ Yes ☐ No ☑ N/A | P-06: complete C2 |
| 2 | 3 Nov 2026, 14:00 | Sponsor asks for the liner lot and the assembly steps (MM-Q-131) | None | MM-Q-131 | 6 Nov 2026 | Answered 5 Nov; EDC receipt | ☑ Yes ☐ No ☐ N/A | P-01: reply with ICDL and explant label |
| 3 | 13 Nov 2026, 16:20 | Sponsor determination: USADE; lot GS-L36-2607 quarantined at all sites | Anticipated status resolved | SAF-GS-0031 | N/A | N/A | ☐ Yes ☐ No ☑ N/A | P-01: quarantine stock; tell the team |
| 4 | 19 Nov 2026, 10:00 | Device analysis: locking ring out of tolerance (DA-0045) | Device cause confirmed; operator contribution reduced | DA-0045 | N/A | N/A | ☐ Yes ☐ No ☑ N/A | P-01: update the SCML context |
| 5 | 9 Dec 2026, 11:00 | 6-week visit after revision: stable, no further dislocation (PFVR-014-005-R6) | Outcome: recovering | N/A | 16 Dec 2026 | Sent 10 Dec; EDC update | ☑ Yes ☐ No ☐ N/A | P-06: 3-month visit, Jan 2027 |
| 6 | | | | | | | ☐ Yes ☐ No ☐ N/A | |
| 7 | | | | | | | ☐ Yes ☐ No ☐ N/A | |
| 8 | | | | | | | ☐ Yes ☐ No ☐ N/A | |

If more lines are needed, reprint this sheet, number it and continue the numbering.

# SART - Safety Assessment, Reporting and Tracking Template

Links and closure

SIMULATED · v1.1

| Study / site | SART case ID / revision | Participant code or N/A | Local closure date |
|---|---|---|---|
| GLENOS-01 / 014 | SART-014-005 / r4 | 014-0005 | Open; next review 15 Jan 2027 |

## F Records this case is sent to

Each line is a record that receives this case from the SART. Mark whether the link was made and write the receiving record's ID.

| Record | When it applies | ID / date | Status |
|---|---|---|---|
| **SCML** | Every adverse event or device deficiency (DD) at the site · SCML event ID | SCML-014-004, 28 Oct 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **Deviation record** | The initial notification was late; the sponsor's criteria decide its importance · Deviation ID | PD-014-12, 1 Nov 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **CSST, early review** | Serious health threat, cluster, new category, or concentration in one component or surgeon · Review reference | Early CSST review, 4 Nov 2026 (serious event, new category) | ☑ Linked ☐ Not required ☐ Pending |
| **POT** | The case, a delay or a pending decision needs the PI's oversight · POT review ID | POT-014-T-2026-11, triggered review | ☑ Linked ☐ Not required ☐ Pending |
| **SRR or CAPA** | A recurring failure or a weak local control; not for a single case · Record ID | CAPA-014-2026-004: escalation while the PI is away | ☑ Linked ☐ Not required ☐ Pending |
| **PIR** | Every reportable case enters the safety reporting timeliness count of its review period · PIR ID / period | PIR-014-2026-11, November 2026 | ☑ Linked ☐ Not required ☐ Pending |
| **Sponsor device-deficiency route** | The device deficiency needs return, analysis or action by the sponsor · Sponsor DD case ID | GS-DD-014-0012; RMA-0045 | ☑ Linked ☐ Not required ☐ Pending |

## G Closing the site's part of the case

| Before closing, confirm that… | Result | Reference / limitation |
|---|---|---|
| Immediate protection and care documented (or N/A, with the reason) | ☑ Yes ☐ No ☐ N/A | Emergency department and revision notes |
| Source references enough to reconstruct the case | ☑ Yes ☐ No ☐ N/A | Section B source map |
| Current PI / site assessment documented | ☑ Yes ☐ No ☐ N/A | C4, 3 Nov 2026 |
| Sponsor submission and receipt documented | ☑ Yes ☐ No ☐ N/A | GS-SAE-014-0003, GS-DD-014-0012 |
| Timeliness and completeness known, or the limitation recorded | ☑ Yes ☐ No ☐ N/A | Late and incomplete: D, PD-014-12 |
| Follow-up completed, transferred or explicitly open | ☑ Yes ☐ No ☐ N/A | Open: 3-month visit, Jan 2027 |
| Sponsor / CEC determination recorded, or recorded as pending | ☑ Yes ☐ No ☐ N/A | SAF-GS-0031, 13 Nov 2026 |
| Links in Section F completed | ☑ Yes ☐ No ☐ N/A | F, all seven linked |

**Local closure status**

☐ Closed locally ☑ Open follow-up ☐ Transferred ☐ Not yet eligible

**Current outcome and open items**

Recovering after revision. Open: sponsor lot investigation and the 3-month visit.

Local closure does not close the sponsor, CEC, ethics committee or regulatory process.

### Case reporter / initial site assessor

Name / role: Dr R. Teixeira / Sub-investigator, PI backup

Signature / date:

R. Teixeira 28 Oct 2026

### PI / accountable investigator

Name / role: Dr Alex Morgan / PI

Signature / date:

A. Morgan 10 Dec 2026

### Safety coordinator (if applicable)

Name / role: Maria Silva / Study Coordinator

Signature / date:

M. Silva 10 Dec 2026

**Controlled location of the SART and linked system references:** Investigator site file, tab 6 (Safety) / SART-014-005; GLENOS EDC GS-SAE-014-0003 and GS-DD-014-0012; SCML-014; RMA-0045