SIRR - Site Imaging Readiness Record

Assessment identification, purpose and local response

SIMULATED · v0.6

Study / protocol GLENOS-01 / GS-2026-01	**CIP version** CIP v4.2	**Site / code** Riverside Orthopedic Center / 014	**SIRR ID / revision** SIRR-014 / r1
Assessment type ☐ Initial ☑ Reassessment	**Assessment date** 2 Jun 2026	**Imaging charter or manual / version / effective date** GLENOS imaging manual IMG-MAN v2.0 / effective 1 Apr 2026	
Charter received / confirmed by / date Received 7 Apr 2026 / Elena Martin / 9 Apr 2026	**Reassessment trigger (if any)** Sponsor reply to Q-IMG-02 (3-month axillary view)	**Radiology contact / role** Elena Martin / Radiology research technologist (P-09)	**Controlled location** Investigator site file, tab 7 (Imaging) / SIRR-014

A Purpose and use

The SIRR records whether the site can carry out the imaging the sponsor's charter requires before the first participant is imaged. Each requirement of the charter is mapped to the local equipment, workflow and people that will deliver it (D to G), gaps get an action (H), and the site concludes whether it can submit its readiness to the sponsor or CRO (I). The charter stays under sponsor control: a requirement the site cannot meet is recorded as a gap or a question, never rewritten.

The SIRR is about the local arrangement, not about any single image set, which is judged in the Image Quality Log (IQL). Signatures attest the site's self-assessment; they do not activate the site, which stays a sponsor decision referenced in K.

Unit of analysis	Decision supported	Inputs	Outputs
The site's imaging arrangement for one version of the sponsor's charter.	Whether the site can submit its readiness for sponsor or CRO review, and with which limits.	CIP, imaging charter or manual, equipment records, transfer test, delegation and training records.	Requirement IDs that the IQL cites, gaps and actions, the site conclusion, and the reassessment history.

B What the site verifies

Each element is assessed in the section shown.

Element	What the site verifies	Section
Acquisition	Modality, views, positioning, parameters, laterality and timing can be delivered locally.	D
Equipment, transfer and archive	Equipment evidence is current; images keep their identifiers through export, transfer, receipt and retrieval.	E
People and review	Acquisition, local review and clinical decision have named, trained people and a handoff to central review.	F, G
Response to a problem image	Who records, repeats, escalates and closes a limited or non-conforming image set.	C, G

C Local response to each image-set category

Use the sponsor's categories if they differ. The IQL applies this response to each image set; the SIRR only defines it.

Category	Local response (who records, repeats, reviews, escalates)	Responsible role
Acceptable Complete, traceable, meets the requirement	Technologist checks views and labels at the console, logs the set in the IQL and sends it through the sponsor portal.	Research technologist
Limited Usable with a stated limitation	Technologist records the limitation in the IQL; local reviewer decides within 2 working days if it can be repeated in window.	Local imaging reviewer
Non-conforming Does not meet the requirement	Repeat in window when possible; if not, IQL limitation and deviation check by the coordinator; PI decides on clinical need.	Local reviewer, PI
Not assessable Set, source or reviewer unavailable	Coordinator recovers the set or its source; if lost, IQL and deviation log; sponsor told within 5 days.	Study coordinator

D Charter requirements and local implementation

One row per requirement. Keep the ID stable: the IQL, PFVR and IMC cite SIRR ID and requirement ID. Copy the requirement as written; a question to the sponsor goes to H. N/A needs the reason in the gap column.

Req. ID	Charter section	Time point · window	Modality · views · positioning · parameters	Local equipment · workflow (ref.)	Implementation	Evidence	Gap · impact on readiness	Action ID (H)
IMG-01	§3.1	Before surgery · ≤ 90 days	CT shoulder, whole scapula; 0.625 mm slices; bone kernel	CT-2 · preset GLENOS-SH · WI-IMG-02	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-02	§3.2	Before surgery · ≤ 30 days	X-ray: true AP (Grashey), axillary, scapular Y	XR-3 · order set SH-3V	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-03	§4.1	Before discharge	X-ray: true AP, axillary	XR-3 or ward unit · order set SH-2V	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-04	§4.2	6 weeks · day 42 ± 7	X-ray: true AP, axillary, scapular Y	XR-3 · order set SH-3V · booking rule BR-06	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-05	§4.3	3 months · day 90 ± 14	X-ray: true AP, axillary, scapular Y	XR-3 · order set SH-3V	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing	Axillary often impossible at 3 months; Velpeau view accepted by sponsor (r1)	A-01
IMG-06	§4.4	12 months · day 365 ± 30	X-ray: true AP, axillary, scapular Y	XR-3 · order set SH-3V	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-07	§4.5	24 months · day 730 ± 30	X-ray: true AP, axillary, scapular Y	XR-3 · order set SH-3V	☑ Ready ☐ Partial ☐ Not ready ☐ N/A	☑ Adequate ☐ Partial ☐ Missing		
IMG-08	§5.1	Unplanned · ≤ 7 days after instability or revision	X-ray: true AP, axillary	Emergency XR-1 · order set SH-2V	☐ Ready ☑ Partial ☐ Not ready ☐ N/A	☐ Adequate ☑ Partial ☐ Missing	Out-of-hours sets not labeled with study ID; observation	A-03

E Equipment, transfer and archive

Cite the current certificate or test and its expiry. A transfer test covers identifiers, export, receipt, local retrieval and retrieval through the sponsor route when it can be tested. No images are copied here.

Control	Equipment or test ID	Current evidence · valid until	Status	Gap · action ID (H)
Equipment and preset (1)	CT-2 (Siemens SOMATOM)	Preset GLENOS-SH checked 20 May 2026 against §3.1	☑ Ready ☐ Partial ☐ Not ready	
Equipment and preset (2)	XR-3 (digital radiography)	Order sets SH-2V and SH-3V live 18 May 2026	☑ Ready ☐ Partial ☐ Not ready	
Maintenance and calibration	CT-2, XR-3, XR-1	Service reports: CT-2 valid to 3 Apr 2027; XR-3 to 12 Mar 2027; XR-1 to 30 Jan 2027	☑ Ready ☐ Partial ☐ Not ready	
Transfer and retrieval test	TT-014-01, TT-014-02	TT-01 failed on CT over 500 slices; TT-02 passed 28 May 2026, receipt and retrieval confirmed by the core lab	☑ Ready ☐ Partial ☐ Not ready	A-02 closed
Local archive and display	Hospital PACS	Retrieval of test set by study ID checked 28 May 2026	☑ Ready ☐ Partial ☐ Not ready	

F People and review boundary

Cite the delegation and training records; this table does not authorize anyone. Sponsor central review and endpoint adjudication keep their own authority.

Function	Name / role	Scope · when to escalate	Delegation · training record	Ready?
Technical acquisition and transfer	Elena Martin / Radiology research technologist	Acquires and checks views; sends sets; escalates a limited set to the reviewer	Delegation log rev. 5; TRE-014-07 (IMG-MAN v2.0)	☑ Yes ☐ No
Local imaging review	Dr Helen Meyer / MSK radiologist	Rates usability and clinical concern; refers disagreement with the core lab to the PI	Delegation log rev. 5; TRE-014-08	☑ Yes ☐ No
Clinical decision and repeat imaging (PI)	Dr Alex Morgan / Principal Investigator	Decides on repeat imaging and clinical action; receives core-lab disagreements	Delegation log rev. 5; GCP certificate 2025	☑ Yes ☐ No

SIRR - Site Imaging Readiness Record

Image-set pathways, gaps and actions

SIMULATED · v0.6

G What happens to an image set

For each situation, name the record or system and who acts. The IQL records each set; the PFVR links a visit to its IQL entry.

Situation	Record or system	Responsible role · handoff	Pathway	Evidence · action ID (H)
Acceptable set sent to central review	IQL; GLENOS imaging portal	Technologist sends within 5 days	☑ Defined and tested ☐ Partial ☐ Not defined	TT-014-02
Limited or non-conforming set	IQL exception; deviation log	Technologist to reviewer; PI if no repeat	☑ Defined and tested ☐ Partial ☐ Not defined	Dry run 30 May 2026
Image needed for eligibility or a clinical decision	PSEAR; PACS report	Reviewer report to PI before surgery	☑ Defined and tested ☐ Partial ☐ Not defined	Dry run 30 May 2026
Repeat acquisition or corrected transfer	IQL action; original and repeat set linked	Technologist; coordinator checks the link	☑ Defined and tested ☐ Partial ☐ Not defined	TT-014-02
Recurrent problem or central-reader disagreement	Core-lab query; SRR, POT or CAPA	PI with the core lab	☐ Defined and tested ☑ Partial ☐ Not defined	Route agreed by email only; A-04

H Gaps, questions and actions

A gap in acquisition, timing, people, transfer or retrieval blocks submission until it is closed. Close an action only with evidence.

Action ID	Req. ID / control	Gap or question to the sponsor	Blocks submission?	Owner (role)	Due date	Status · closing evidence
A-01	IMG-05	Q-IMG-02 to sponsor: is a Velpeau axillary acceptable at 3 months?	☐ Yes ☑ No	Technologist	30 May 2026	Closed: sponsor reply SP-IMG-0214, 31 May 2026
A-02	Transfer	CT over 500 slices failed transfer test TT-014-01	☑ Yes ☐ No	Technologist	28 May 2026	Closed: TT-014-02 passed
A-03	IMG-08	Out-of-hours X-ray not labeled with study ID	☐ Yes ☑ No	Coordinator	30 Jun 2026	Open: label request to radiology
A-04	G, row 5	Core-lab disagreement route only agreed by email	☐ Yes ☑ No	PI	15 Jun 2026	Open: written in site imaging instruction
			☐ Yes ☐ No			
			☐ Yes ☐ No			

SIRR - Site Imaging Readiness Record

Conclusion, history and attestation

SIMULATED · v0.6

I Site conclusion

Select one for this version of the charter.

Conclusion	Meaning	Select
Ready for sponsor or CRO review	Every requirement is mapped and supported by current evidence.	☐
Ready, with a documented observation	An open point that does not block any requirement; it has an owner, a due date and is disclosed to the sponsor.	☑
Action required before submission	A gap in H blocks submission.	☐
Not ready: sponsor clarification or escalation	The site cannot meet a requirement as written.	☐
Open observations (action IDs)	A-03 (labels out of hours), A-04 (disagreement route): disclosed in the submission letter	

J Assessment history

Reopen the SIRR after a charter or CIP change, new equipment, software or archive, failed maintenance, a change of acquisition staff or reviewer, positioning that cannot be reproduced, or a recurrent problem in the IQL or an IMC finding. Keep earlier rows.

Rev.	Date	Trigger · charter version · requirements affected	Conclusion	Next review or trigger
r0	20 May 2026	Initial · IMG-MAN v2.0 · all	Action required before submission (A-01, A-02)	After A-01 and A-02
r1	2 Jun 2026	Sponsor reply SP-IMG-0214; TT-014-02 · v2.0 · IMG-05, transfer	Ready, with a documented observation	Charter amendment or IQL signal

K Attestation and sponsor references

Copy sponsor or CRO references from their controlled records. Activation is the sponsor's decision.

Submitted to sponsor or CRO (reference / date)	SIRR-014 r1 package, sent to the GLENOS core lab / 3 Jun 2026
Sponsor or CRO response (reference / date)	Core lab qualification letter IQ-014 / 9 Jun 2026: site qualified for imaging
Site activation record (reference / date)	Sponsor activation letter GS-ACT-014 / 16 Jun 2026

Technical acquisition

Name / role: E. Martin / P-09

Signature / date:

E. Martin 2 Jun 2026

Local imaging reviewer

Name / role: Dr H. Meyer / MSK

Signature / date:

H. Meyer 2 Jun 2026

Principal Investigator

Name / role: Dr A. Morgan / PI

Signature / date:

A. Morgan 3 Jun 2026