SART - Safety Assessment, Reporting and Tracking Template

Case opening, purpose and awareness

SART · v1.1

Study / protocol	**Current CIP / version**	**Site / code**	**SART case ID / revision**
Participant code (or N/A and why)	**Procedure, visit or exposure (or N/A)**	**Site awareness: date / time**	**Person who received the first information**
PI / authorized site assessor	**Sponsor safety system / route**	**Sponsor safety case ID (if issued)**	**Current local case status** ☐ Open ☐ Awaiting sponsor ☐ Closed locally

A Purpose and use

The SART is the site's record of one safety case, from the moment the site becomes aware of it until the site's part of the case is closed. It shows that the site recognized the information, protected the person, made its own assessment, reported through the sponsor's route within the CIP window, followed the case and linked it to its other records.

Open the SART at awareness, without waiting for a complete classification. Sections B to D are completed in the first days; Section E grows with each new piece of information; Sections F and G close the site's part.

Boundary. This is a controlled site coordination record. It does not replace source records, the sponsor safety form, the eCRF, the sponsor safety portal, the device-accountability record, the deviation record, the SCML, the CSST, the CAPA log, a regulatory submission or the sponsor/CEC safety determination. It records the site's part of the case and the references to those records.

Unit of analysis	One safety case at the site: one adverse event, one observed device deficiency or one piece of new safety information.
Decision supported	Whether the site recognized, protected, assessed, reported, followed and closed its part of the case in a controlled way.
Inputs	Source records, participant or user contact, device and accountability records, procedure or visit records (ICDL, PFVR, IQL), sponsor communications, the current CIP and risk information.
Outputs	The local trail of the case; references to the sponsor system; the reporting data read by the PIR (safety reporting timeliness); the links to the SCML, the deviation record and, when triggered, the CSST, POT, SRR or CAPA.

Rule: if an adverse event is associated with an observed device deficiency, apply both the adverse-event pathway and the device-deficiency pathway (Section C).

B Awareness, immediate protection and source map

Record what is known at awareness and where it is documented. Do not copy narratives from the clinical record or the sponsor form.

Item	Status	Record / reference
Occurrence, deficiency detection or new information: date / time	☐ Known ☐ Estimated	
Site awareness: date / time, and who brought the information (person and source)		
Initial trigger	☐ Adverse event ☐ Observed device deficiency ☐ AE associated with an observed DD ☐ New safety information ☐ Other CIP-defined trigger	
Immediate safety concern for the participant, user or another person	☐ Yes ☐ No ☐ Not yet determined	
Immediate action taken: clinical care, device containment or communication (who, when)		
Device and procedure context: system and component; model, lot, serial or UDI; size; technical step (or N/A)		

SART - Safety Assessment, Reporting and Tracking Template

Awareness and immediate protection

Study / site	SART case ID / revision	Participant code or N/A	Site awareness: date / time

B Awareness, immediate protection and source map (continued)

Factual summary (what happened, as known now)

Source map

Source record	ID / location	What it holds for this case

Missing information

Information still missing	Owner (Role ID)	Next step / due

SART - Safety Assessment, Reporting and Tracking Template

Current PI / site assessment

Study / site	SART case ID / revision	Participant code or N/A	Assessment date

C Current PI / site assessment

This is the site's assessment and may change. The sponsor or CEC determination is recorded apart, in the last table. Rationale and source go in the right-hand column.

C1 · Adverse-event pathway		
Adverse event identified?	☐ Yes ☐ No ☐ Pending	
Seriousness (mark all that apply)	☐ Death ☐ Life-threatening illness or injury ☐ Permanent impairment of a body structure or function ☐ In-patient or prolonged hospitalization ☐ Intervention to prevent the above ☐ Fetal distress, fetal death or congenital abnormality ☐ Does not meet seriousness criteria ☐ Pending	
Clinical severity (intensity, on the CIP scale): kept apart from seriousness	☐ Mild ☐ Moderate ☐ Severe ☐ Pending	
Complication grade (Clavien-Dindo), by the treatment the event needed	☐ I ☐ II ☐ IIIa ☐ IIIb ☐ IVa ☐ IVb ☐ V ☐ N/A ☐ Pending	
Related to the investigational device?	☐ Yes ☐ No ☐ Pending	
Related to the comparator?	☐ Yes ☐ No ☐ Pending ☐ N/A	
Related to a CIP-required procedure outside routine practice?	☐ Yes ☐ No ☐ Pending ☐ N/A	
If a SADE: identified in the current risk assessment?	☐ ASADE ☐ USADE ☐ Pending ☐ N/A	
Resulting local category	☐ Non-serious AE ☐ SAE ☐ ADE ☐ SADE ☐ ASADE ☐ USADE ☐ Pending	

C2 · Device-deficiency pathway		
Observed device deficiency?	☐ Yes ☐ No ☐ Pending	
Type (mark all that apply)	☐ Identity ☐ Quality ☐ Durability ☐ Reliability ☐ Usability ☐ Safety ☐ Performance ☐ Malfunction ☐ Use error ☐ Information or label	
Associated with an adverse event? (Yes: apply both pathways)	☐ Yes ☐ No ☐ Pending	
Could it have led to a SADE?	☐ Yes ☐ No ☐ Pending	
Device containment, return or analysis (reference)	☐ Done ☐ Pending ☐ N/A	

C3 · Context and urgency		
Contributing context, as now assessed (for the SCML; mark all that apply)	☐ Device ☐ Procedure ☐ Operator or technique ☐ Patient ☐ Undetermined	
Serious health threat (imminent risk needing rapid action to protect others)?	☐ Yes ☐ No ☐ Pending	

C4 · Who has assessed the case		
Current PI / site assessment: status, assessor, date	☐ Pending ☐ Preliminary ☐ Final (PI)	
Sponsor / CEC determination: date and reference	☐ Not yet received ☐ Received	
Agreement between the two	☐ Agree ☐ Disagree ☐ Not yet known	
Open classification question and owner		

SART - Safety Assessment, Reporting and Tracking Template

Reporting to the sponsor

Study / site	SART case ID / revision	Participant code or N/A	Sponsor safety case ID

D Reporting to the sponsor

Record the proof of reporting, not the content of the sponsor form. Every time limit comes from the CIP or the sponsor procedure. The PIR reads its safety reporting timeliness counts from this section.

Reporting step	Status	Record / reference
Reportable under the current CIP or sponsor procedure? (rule reference)	☐ Yes ☐ No ☐ Pending	
Route used	☐ Sponsor portal ☐ eCRF ☐ SAE form ☐ DD form ☐ Emergency route ☐ Other	
Site awareness: date / time (from Section B)		
CIP reporting window, and due date / time		
Initial notification sent: date / time, by whom, system reference		
Receipt confirmed / sponsor safety case ID	☐ Yes ☐ Pending	
Initial notification on time?	☐ Yes ☐ No ☐ Pending ☐ N/A	
First report complete?	☐ Complete ☐ Incomplete ☐ Pending ☐ N/A	
If late or incomplete: where the delay arose	☐ Recognition ☐ Internal escalation ☐ PI assessment ☐ Access to documentation ☐ Transmission ☐ Other	
Correction made, and deviation record ID if the notification was late		
Ethics committee or authority communication, only if the CIP assigns it to the site	☐ Sent ☐ Not required ☐ Pending	

SART - Safety Assessment, Reporting and Tracking Template

Follow-up and new information

Study / site	SART case ID / revision	Participant code or N/A	Sheet number / total sheets

E Follow-up and new information

One line for each new piece of information or sponsor request. Add lines; never overwrite an earlier one.

No.	Date / time received	New clinical, technical or device information, and its source	Effect on the site assessment	Sponsor request or communication (reference)	Follow-up due	Sent / receipt reference	On time?	Owner (Role ID) and next step
1							☐ Yes ☐ No ☐ N/A	
2							☐ Yes ☐ No ☐ N/A	
3							☐ Yes ☐ No ☐ N/A	
4							☐ Yes ☐ No ☐ N/A	
5							☐ Yes ☐ No ☐ N/A	
6							☐ Yes ☐ No ☐ N/A	
7							☐ Yes ☐ No ☐ N/A	
8							☐ Yes ☐ No ☐ N/A	

If more lines are needed, reprint this sheet, number it and continue the numbering.

SART - Safety Assessment, Reporting and Tracking Template

Links and closure

Study / site	SART case ID / revision	Participant code or N/A	Local closure date

F Records this case is sent to

Each line is a record that receives this case from the SART. Mark whether the link was made and write the receiving record's ID.

Record	When it applies	ID / date	Status
SCML	Every adverse event or device deficiency (DD) at the site · SCML event ID		☐ Linked ☐ Not required ☐ Pending
Deviation record	The initial notification was late; the sponsor's criteria decide its importance · Deviation ID		☐ Linked ☐ Not required ☐ Pending
CSST, early review	Serious health threat, cluster, new category, or concentration in one component or surgeon · Review reference		☐ Linked ☐ Not required ☐ Pending
POT	The case, a delay or a pending decision needs the PI's oversight · POT review ID		☐ Linked ☐ Not required ☐ Pending
SRR or CAPA	A recurring failure or a weak local control; not for a single case · Record ID		☐ Linked ☐ Not required ☐ Pending
PIR	Every reportable case enters the safety reporting timeliness count of its review period · PIR ID / period		☐ Linked ☐ Not required ☐ Pending
Sponsor device-deficiency route	The device deficiency needs return, analysis or action by the sponsor · Sponsor DD case ID		☐ Linked ☐ Not required ☐ Pending

G Closing the site's part of the case

Before closing, confirm that…	Result	Reference / limitation
Immediate protection and care documented (or N/A, with the reason)	☐ Yes ☐ No ☐ N/A	
Source references enough to reconstruct the case	☐ Yes ☐ No ☐ N/A	
Current PI / site assessment documented	☐ Yes ☐ No ☐ N/A	
Sponsor submission and receipt documented	☐ Yes ☐ No ☐ N/A	
Timeliness and completeness known, or the limitation recorded	☐ Yes ☐ No ☐ N/A	
Follow-up completed, transferred or explicitly open	☐ Yes ☐ No ☐ N/A	
Sponsor / CEC determination recorded, or recorded as pending	☐ Yes ☐ No ☐ N/A	
Links in Section F completed	☐ Yes ☐ No ☐ N/A	

Local closure status	Current outcome and open items
☐ Closed locally ☐ Open follow-up ☐ Transferred ☐ Not yet eligible	

Local closure does not close the sponsor, CEC, ethics committee or regulatory process.

Case reporter / initial site assessor

Name / role:

Signature / date:

PI / accountable investigator

Name / role:

Signature / date:

Safety coordinator (if applicable)

Name / role:

Signature / date:

Controlled location of the SART and linked system references: