ICDL — INTRAOPERATIVE CHECKLIST AND DEVICE LOG

ICDL · v0.12

Study / Protocol	Site / Code	Participant / Case Code	ICDL Record ID
Procedure Date	Procedure / Side	Operating Room	Primary Surgeon
Device / System	Current CIP / Procedure Instructions / Version		Accountability File / Record Location

A Scope, decision and interfaces

Purpose and use: Complete this controlled record during the procedure, not retrospectively. It captures critical procedural execution, unit-level device traceability, intraoperative decisions and immediate post-procedure review at the point of generation. It supports, but does not replace, the operative source note, authoritative device-accountability record, safety reporting, deviation documentation, imaging source records or CAPA.

Element	Complete when	Record / link
Unit of analysis	Open one ICDL for one participant and one procedure.	Participant code, procedure date and ICDL ID are unique and reconciled.
Decision supported	Close only when critical steps, device identity, deviations and immediate outcome are documented or escalated.	Case-closure status and unresolved route.
Inputs	Approved plan / CIP, credentialing / TRE release, device instructions and accountability materials.	Reference IDs or locations; do not copy full narratives.
Outputs	Contemporaneous checklist, device traceability, deviation/substitution assessment and immediate review.	Operative note, device accountability record, SCML/SART, ICQL, deviation or CAPA as applicable.

B Pre-procedure Release-to-Start and Critical Controls

Readiness or critical control	Status	Evidence / current reference	Exception, hold or immediate route
Authorised surgeon and current release confirmed	☐ Yes ☐ No ☐ N-A		
Participant identity, procedure and side match the approved schedule and source records	☐ Yes ☐ No ☐ N-A		
Required consent, eligibility and pre-procedure source checks are available	☐ Yes ☐ No ☐ N-A		
Device system, required components and instruments are available and compatible	☐ Yes ☐ No ☐ N-A		
Packaging, expiry, storage and sterile release conditions are acceptable	☐ Yes ☐ No ☐ N-A		
Required image, measurement or case-specific materials are available	☐ Yes ☐ No ☐ N-A		

Release-to-start decision	Decision time	Responsible reviewer	Linked hold / action record
☐ Released ☐ Released with observation ☐ Hold / action required			

C Critical Technique Checklist

Complete one row for each critical technique step defined by the CIP, procedural plan, device instructions or site-approved procedure checklist. Do not reproduce every operative action. Where a step is omitted, adapted, repeated or performed outside the expected sequence, preserve the fact in this record and continue to Section F.

No.	Critical technique step / control point	Execution status	Evidence / source reference	Immediate result	Owner / initials
Example	*Final component orientation confirmed against the approved plan before closure.*	☑ *Complete* ☐ *Partial* ☐ *Not performed* ☐ *N/A*	*Operative note 14:35; image ref. IMG-02.*	☑ *Acceptable* ☐ *Observation* ☐ *Not acceptable* ☐ *Not assessable*	*AB*
1		☐ Complete ☐ Partial ☐ Not performed ☐ N/A		☐ Acceptable ☐ Observation ☐ Not acceptable ☐ Not assessable	
2		☐ Complete ☐ Partial ☐ Not performed ☐ N/A		☐ Acceptable ☐ Observation ☐ Not acceptable ☐ Not assessable	
3		☐ Complete ☐ Partial ☐ Not performed ☐ N/A		☐ Acceptable ☐ Observation ☐ Not acceptable ☐ Not assessable	
4		☐ Complete ☐ Partial ☐ Not performed ☐ N/A		☐ Acceptable ☐ Observation ☐ Not acceptable ☐ Not assessable	
5		☐ Complete ☐ Partial ☐ Not performed ☐ N/A		☐ Acceptable ☐ Observation ☐ Not acceptable ☐ Not assessable	

D Procedure Time Points and Operating Conditions

Time point	Time	Condition / status	Reference / initials
Procedure start / first incision			
Device opened / first component used			
Final implant / configuration confirmed			
Closure / procedure end			

E Device Accountability and Component-Level Traceability

Record every component used, including auxiliary, replacement, trial, explanted or discarded components where applicable. Copy identifiers from the label or authorised source at the time of use. Do not infer missing identifiers later. Reconcile the final configuration against the operative note and the authoritative device-accountability record before case closure. If a component or final configuration cannot be reconciled, do not close the case as complete; record the linked route in Section F.

No.	Component / role	Manufacturer / system	Model / catalogue	UDI-DI / device ID	UDI-PI / lot / serial	Size / config.	Disposition	Source ref.
Example	*Humeral component*	*Sponsor / system A*	*HC-42*	*123456789*	*LOT-2026-04 / SN-18*	*42 mm*	☑ *Implanted* ☐ *Trial* ☐ *Returned* ☐ *Discarded* ☐ *Explanted*	*Label 01*
1							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	
2							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	
3							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	
4							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	
5							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	
6							☐ Implanted ☐ Trial ☐ Returned ☐ Discarded ☐ Explanted	

FINAL IMPLANTED CONFIGURATION

Components, sizes, orientation and other configuration details required by the current procedure.

ACCOUNTABILITY RECONCILIATION REFERENCE

Device log, operative note, label set or other authoritative reference used for reconciliation.

F Technical Deviation, Substitution and Exception Route

Open this section whenever the planned technique, component, operating condition or approved sequence is changed, omitted, repeated or made uncertain. The entry must preserve what happened, why it happened, the potential impact and the immediate route; it does not replace the authoritative deviation record where one is required.

No.	Signal / exception	What happened and when	Reason / technical context	Immediate impact assessment	Immediate response / linked record
Example	*Component substitution*	*Glenoid size changed at 14:12 after trial fit.*	*Unexpected morphology.*	*Assess endpoint and safety impact; no immediate instability.*	*PI informed; DEV-003; operative note ref.*
1					
2					
3					
4					

G Intraoperative Complications and Device Performance

No.	Event / performance concern	Clinical description	Device or procedure relationship	Treatment / resolution at case	Linked safety record
Example	*Instrument malfunction*	*Locking mechanism failed during trial.*	*Possible device deficiency; no participant harm observed.*	*Instrument replaced; procedure completed.*	*SCML-004 / DD-002*
1					
2					
3					
4					

Immediate safety rule: this ICDL is not the place to wait for aggregate review. Any event, deficiency or uncertainty that may affect participant safety, device accountability or endpoint interpretation must be routed immediately through the applicable local and sponsor pathway.

H Immediate Post-procedure Review

Complete before the case is closed by the site. The review confirms whether the contemporaneous record is usable, the final configuration is reconciled, the immediate postoperative status is documented and all required imaging, deviation, device-deficiency and safety routes are linked.

Review item	Status	Evidence / reference	Open action / owner
Operative note contains final configuration, critical decisions, substitutions, deviations and complications	☐ Adequate ☐ Partial ☐ Not adequate ☐ N/A		
Device component list is complete and reconciled with labels and accountability record	☐ Adequate ☐ Partial ☐ Not adequate ☐ N/A		
Immediate postoperative condition and clinically relevant findings are documented in source records	☐ Adequate ☐ Partial ☐ Not adequate ☐ N/A		
Required immediate postoperative image or image reference has been routed to ICQL / source record	☐ Adequate ☐ Partial ☐ Not adequate ☐ N/A		
All safety, deviation and device-deficiency signals have a linked route	☐ Adequate ☐ Partial ☐ Not adequate ☐ N/A		

IMMEDIATE POSTOPERATIVE REVIEW SUMMARY

IMAGE / SOURCE / RECORD REFERENCES

I Case Closure Decision

Closure status	Closure time	Responsible reviewer	Linked open action / record
☐ Complete ☐ Complete with open action ☐ Hold			

Complete means the record is reconstructable and all required routes are linked. **Complete with open action** requires a named owner, due date and destination. **Hold** means the case must not be treated as closed until the blocker is resolved or formally escalated.

J Open Action, Escalation and Effectiveness Follow-up

Use this section only for an unresolved issue, limitation, signal or action arising from Sections B–I. The receiving record remains authoritative for its own content. This table preserves the hand-off, ownership, due date and evidence that the local response was effective.

No.	Issue / linked section	Destination record / escalation	Owner	Due date	Effectiveness evidence	Status / closure note
Example	*F — component substitution*	*DEV-003 / operative note / PI review*	*AB*	*Same day*	*Final configuration reconciled and no further action required.*	☐ *Open* ☑ *Closed*
1						☐ Open ☐ Closed
2						☐ Open ☐ Closed
3						☐ Open ☐ Closed
4						☐ Open ☐ Closed

K Accountable Review, Approval and Controlled Record Location

Completed by — intraoperative record owner

Name / role _____

Signature / initials _____

Date / time _____

Surgeon / authorised procedure reviewer

Name / role _____

Signature / initials _____

Date / time _____

PI / accountable reviewer, where required

Name / role _____

Signature _____

Date _____

Controlled record location and retained links

ICDL location _____

Accountability record _____

Safety / deviation / imaging _____

System linkage: Route an individual complication or device-performance concern to the SCML and the applicable safety process; an image or transfer issue to the ICQL; a recurring process-control failure to the IMC, SRR or CAPA pathway; and a formal protocol deviation to the record defined by the CIP.