IQL — IMAGING QUALITY LOG

Purpose, scope and controlled use

Study / Protocol	Site / Code	IQL Record ID	Review period / cut-off date
Applicable Imaging Charter ID / version	**Modality / time-point scope included**	**Prepared by / date**	**Controlled record location**

A Purpose, Scope, Decision and Use

Purpose and use: Use one IQL for a defined site, protocol, applicable sponsor Imaging Charter version, review period and imaging scope. It records whether each required image set was acquired within the permitted window, technically acceptable and complete, traceable through archive or transfer, and locally reviewed; it also documents whether an exception requires correction, repeat acquisition, second review or escalation. The IQL supports site-level imaging quality control against sponsor-defined requirements. It does not replace the source images, radiology report, sponsor or central review, endpoint adjudication, safety reporting, protocol-deviation documentation or sponsor oversight.

Element	Operational definition / complete when	Evidence, decision or record link
Unit of analysis	One Section B row for one required image set, for one participant or case and one protocol time point.	Participant or case code, time point, acquisition date and unique archive or transfer reference.
Scope	Only modalities, views, sequences and time points required by the applicable sponsor Imaging Charter or imaging manual for the stated period.	Charter or manual ID and version, review-period dates and included modality or time-point scope.
Decision supported	Determine timing status, technical acceptability, completeness and traceability, and whether repeat, second review, correction or escalation is required.	Section B status, exception or action reference, reviewer and review date.
Inputs	Protocol or CIP, sponsor imaging requirements, visit schedule, acquisition metadata, archive or transfer evidence and authorised local review records.	Reference IDs or controlled locations; do not copy full clinical narratives or image interpretations.
Outputs	Reconstructable image-set status, exception response, denominator-based periodic review, accountable action and current quality-control conclusion.	Sections B–F, linked deviation, PFVR, safety, CAPA, central-review or other governed record as applicable.
Boundary and ownership	The site documents local execution and quality control. Sponsor, central reviewer and adjudication responsibilities remain in their authoritative systems.	Escalate or cross-reference the governed record; do not use the IQL to replace sponsor decisions or clinical source documentation.

Completion sequence: define the controlled scope and review period on this page; add Section B rows as image sets become available; complete Sections C–D at the scheduled review or an immediate trigger; then document the current conclusion and accountable review in Sections E–F.

B Image-Set Quality Log Against Sponsor Imaging Requirements

One row represents one required image set for one participant at one protocol time point. Assess the set against the applicable sponsor Imaging Charter or imaging manual. Record timing, required views or sequences, technical quality, completeness, transfer or archive traceability, local review and any linked response. Do not reproduce the clinical interpretation or image data held in the authoritative records.

Log ID	Participant / case code	Time point	Modality / views or sequences	Planned / actual date	Timing status	Technical quality and completeness	Archive / transfer reference	Reviewer / date	Exception / action / linked record
01					☐ Within window ☐ Outside window ☐ Not applicable	☐ Acceptable ☐ Limited ☐ Non-conforming			
02					☐ Within window ☐ Outside window ☐ Not applicable	☐ Acceptable ☐ Limited ☐ Non-conforming			
03					☐ Within window ☐ Outside window ☐ Not applicable	☐ Acceptable ☐ Limited ☐ Non-conforming			
04					☐ Within window ☐ Outside window ☐ Not applicable	☐ Acceptable ☐ Limited ☐ Non-conforming			
05					☐ Within window ☐ Outside window ☐ Not applicable	☐ Acceptable ☐ Limited ☐ Non-conforming			

A non-conforming set should state whether repeat acquisition is possible, whether the limitation affects an imaging-dependent decision and which linked record documents the response. Attach continuation pages in the same column order when needed.

C Periodic IQL Review and Imaging-Quality Signal Assessment

Summarise the Section B rows for the stated review period at the cadence defined in the applicable sponsor imaging requirements, and perform an additional review when an exception could affect eligibility, procedure documentation, safety evaluation or endpoint interpretation. State each denominator, preserve modality and time-point distinctions unless a documented rationale supports aggregation, and link each decision or follow-up record.

Review measure or control domain	Current period count / denominator	Previous period / comparator	Interpretation	Decision / linked record
Required image sets obtained within the applicable window			☐ Complete ☐ Partial ☐ Not assessable	
Image sets technically acceptable and complete			☐ Stable ☐ Improving ☐ Deteriorating ☐ Not assessable	
Image sets with transfer, archive or naming discrepancy			☐ Isolated ☐ Recurrent ☐ Concentrated ☐ Not assessable	
Image sets requiring repeat acquisition, second review or clinical escalation			☐ Expected ☐ Observation ☐ Action required	
Equipment, workflow or personnel control status			☐ Effective ☐ Limitation ☐ Not effective ☐ Not assessable	

D Exception, Action and Effectiveness Record

Signal, exception or review reference	Immediate response / impact	Corrective or preventive action	Owner / due date	Effectiveness evidence / linked closure

Use a linked record for formal protocol deviations, safety events, study-wide central review findings or actions governed elsewhere. An imaging-quality action is closed only when the required evidence shows that the corrected process is usable for the affected scope.

E Current Period Quality-Control Conclusion

Base this conclusion on the image sets included in the stated review period and on the applicable Imaging Charter version. Record participant-level clinical or safety decisions in the authoritative source and safety records rather than in this summary.

Current conclusion	Meaning and response	Conclusion date / next review
Controlled / no action	☐ The reviewed imaging scope is technically controlled and no additional quality action is required.	
Controlled with observation	☐ An observation is being monitored and does not presently prevent the defined imaging scope.	
Action required	☐ A quality signal or non-conformity requires documented correction and effectiveness follow-up.	
Restricted / escalated	☐ The issue affects use of imaging, an imaging-dependent decision or a formally governed escalation route.	

F Periodic Review, Approval and Record Location

Quality-log preparer / radiology liaison

Name / role

Signature / date

Local clinical imaging reviewer

Name / role

Signature / date

Principal Investigator / accountable reviewer

Name

Signature / date

Next review and controlled record location

Next review / cut-off date

Record location / IQL version

Record linkage: retain the applicable Imaging Charter ID and version in the IQL header. Link participant-level exceptions to the relevant source, PFVR, deviation or safety record; link recurrent site-level control failures to the applicable IMC, SRR or CAPA pathway.