SART - Safety Assessment, Reporting and Tracking Template

Case opening, purpose and awareness

SIMULATED · v1.0

Study / protocol GLENOS-01 / GS-2026-01	**Current CIP / version** CIP v4.2	**Site / code** Riverside Orthopedic Center / 014	**SART case ID / revision** SART-014-005 / r4
Participant code (or N/A and why) 014-0005	**Procedure, visit or exposure (or N/A)** Reverse shoulder arthroplasty, 9 Oct 2026 (ICDL-014-005); revision 28 Oct 2026	**Site awareness: date / time** 28 Oct 2026, 08:05	**Person who received the first information** Maria Silva (P-06), from the hospital admission list
PI / authorized site assessor Dr Alex Morgan (on leave 27 Oct to 2 Nov); backup Dr R. Teixeira	**Sponsor safety system / route** GLENOS EDC: SAE form and DD form	**Sponsor safety case ID (if issued)** GS-SAE-014-0003	**Current local case status** ☑ Open ☐ Awaiting sponsor ☐ Closed locally

A Purpose and use

The SART is the site's record of one safety case, from the moment the site becomes aware of it until the site's part of the case is closed. It shows that the site recognized the information, protected the person, made its own assessment, reported through the sponsor's route within the CIP window, followed the case and linked it to its other records.

Open the SART at awareness, without waiting for a complete classification. Sections B to D are completed in the first days; Section E grows with each new piece of information; Sections F and G close the site's part.

Boundary. This is a controlled site coordination record. It does not replace source records, the sponsor safety form, the eCRF, the sponsor safety portal, the device-accountability record, the deviation record, the SCML, the CSST, the CAPA log, a regulatory submission or the sponsor/CEC safety determination. It records the site's part of the case and the references to those records.

Unit of analysis	One safety case at the site: one adverse event, one observed device deficiency or one piece of new safety information.
Decision supported	Whether the site recognized, protected, assessed, reported, followed and closed its part of the case in a controlled way.
Inputs	Source records, participant or user contact, device and accountability records, procedure or visit records (ICDL, PFVR, IQL), sponsor communications, the current CIP and risk information.
Outputs	The local trail of the case; references to the sponsor system; the reporting data read by the PIR (safety reporting timeliness); the links to the SCML, the deviation record and, when triggered, the CSST, POT, SRR or CAPA.

Rule: if an adverse event is associated with an observed device deficiency, apply both the adverse-event pathway and the device-deficiency pathway (Section C).

B Awareness, immediate protection and source map

Record what is known at awareness and where it is documented. Do not copy narratives from the clinical record or the sponsor form.

Item	Status	Record / reference
Occurrence, deficiency detection or new information: date / time	☑ Known ☐ Estimated	27 Oct 2026, 21:30, at home, while dressing (day 18)
Site awareness: date / time, and who brought the information (person and source)	28 Oct 2026, 08:05; Maria Silva (P-06), from the hospital admission list; emergency department note	
Initial trigger	☑ Adverse event ☐ Observed device deficiency ☐ AE associated with an observed DD ☐ New safety information ☐ Other CIP-defined trigger	Dislocation. The device deficiency was found at revision (Section E, line 1)
Immediate safety concern for the participant, user or another person	☑ Yes ☐ No ☐ Not yet determined	Dislocated shoulder; participant admitted
Immediate action taken: clinical care, device containment or communication (who, when)	Closed reduction failed in the emergency department; admitted. Backup PI Dr R. Teixeira informed 28 Oct, 08:40; open revision 28 Oct, 14:00	
Device and procedure context: system and component; model, lot, serial or UDI; size; technical step (or N/A)	GLENOS v2; humeral liner 36 mm +3, lot GS-L36-2607 (read from the explant label: missing in ICDL-014-005); assembly of the liner on the stem	

SART - Safety Assessment, Reporting and Tracking Template

Awareness and immediate protection

SIMULATED · v1.0

Study / site	SART case ID / revision	Participant code or N/A	Site awareness: date / time
GLENOS-01 / 014	SART-014-005 / r4	014-0005	28 Oct 2026, 08:05

B Awareness, immediate protection and source map (continued)

Factual summary (what happened, as known now)

Eighteen days after a reverse shoulder arthroplasty, the participant felt the shoulder give way while dressing and came to the emergency department, where closed reduction failed. At open revision on 28 Oct the humeral liner was found separated from the stem, with the locking ring not engaged. The liner was replaced, the shoulder was stable at closure, and the removed liner was kept for return to the sponsor.

Source map

Source record	ID / location	What it holds for this case
Emergency department note and admission	EHR, 28 Oct 2026	Presentation, failed reduction, admission
Revision operative note	EHR / ICDL-014-005-R	Liner separated; locking ring not engaged; liner replaced
Explant and return record	RMA-0045, 30 Oct 2026	Liner kept and returned to the sponsor
Index procedure log	ICDL-014-005, Section E	Liner lot not recorded at the index procedure
Sponsor SAE and DD forms	GS-SAE-014-0003; GS-DD-014-0012	Formal report and follow-up
Device analysis report	DA-0045, 19 Nov 2026	Locking ring out of tolerance in lot GS-L36-2607

Missing information

Information still missing	Owner (Role ID)	Next step / due
Sponsor lot investigation outcome	P-01	Sponsor; ask at the next contact
3-month visit after revision	P-06	Visit window 20 Jan to 3 Feb 2027

SART - Safety Assessment, Reporting and Tracking Template

Current PI / site assessment

SIMULATED · v1.0

Study / site	SART case ID / revision	Participant code or N/A	Assessment date
GLENOS-01 / 014	SART-014-005 / r4	014-0005	28 Oct 2026; updated 13 Nov

C Current PI / site assessment

This is the site's assessment and may change. The sponsor or CEC determination is recorded apart, in the last table. Rationale and source go in the right-hand column.

C1 · Adverse-event pathway		
Adverse event identified?	☑ Yes ☐ No ☐ Pending	Emergency department note, 28 Oct
Seriousness (mark all that apply)	☐ Death ☐ Life-threatening illness or injury ☐ Permanent impairment of a body structure or function ☑ In-patient or prolonged hospitalization ☐ Intervention to prevent the above ☐ Fetal distress, fetal death or congenital abnormality ☐ Does not meet seriousness criteria ☐ Pending	Admitted for revision surgery
Clinical severity (intensity, on the CIP scale): kept apart from seriousness	☐ Mild ☐ Moderate ☑ Severe ☐ Pending	CIP v4.2 intensity scale
Related to the investigational device?	☑ Yes ☐ No ☐ Pending	Liner separation caused the dislocation
Related to the comparator?	☐ Yes ☐ No ☐ Pending ☑ N/A	Single-arm study
Related to a CIP-required procedure outside routine practice?	☐ Yes ☑ No ☐ Pending ☐ N/A	Event linked to the device, not to an added procedure
If a SADE: identified in the current risk assessment?	☐ ASADE ☐ USADE ☑ Pending ☐ N/A	Dislocation is in the risk assessment; liner separation is not
Resulting local category	☐ Non-serious AE ☐ SAE ☐ ADE ☑ SADE ☐ ASADE ☐ USADE ☐ Pending	Anticipated status for the sponsor

C2 · Device-deficiency pathway		
Observed device deficiency?	☑ Yes ☐ No ☐ Pending	Found at revision
Type (mark all that apply)	☐ Identity ☐ Quality ☐ Durability ☐ Reliability ☐ Usability ☐ Safety ☑ Performance ☑ Malfunction ☐ Use error ☐ Information or label	Locking ring did not hold the liner
Associated with an adverse event? (Yes: apply both pathways)	☑ Yes ☐ No ☐ Pending	Both pathways apply
Could it have led to a SADE?	☑ Yes ☐ No ☐ Pending	It did lead to one
Device containment, return or analysis (reference)	☑ Done ☐ Pending ☐ N/A	Liner returned, RMA-0045

C3 · Context and urgency		
Contributing context, as now assessed (for the SCML; mark all that apply)	☑ Device ☐ Procedure ☑ Operator or technique ☐ Patient ☐ Undetermined	Liner seating not documented at the index procedure
Serious health threat (imminent risk needing rapid action to protect others)?	☐ Yes ☐ No ☑ Pending	Sponsor checking lot GS-L36-2607 at other sites

C4 · Who has assessed the case		
Current PI / site assessment: status, assessor, date	☐ Pending ☐ Preliminary ☑ Final (PI)	A. Morgan, 3 Nov 2026
Sponsor / CEC determination: date and reference	☐ Not yet received ☑ Received	13 Nov 2026, SAF-GS-0031: USADE, device-related
Agreement between the two	☑ Agree ☐ Disagree ☐ Not yet known	PI had left the anticipated status to the sponsor
Open classification question and owner	Operator contribution to liner seating; P-01, after the device analysis	

SART - Safety Assessment, Reporting and Tracking Template

Reporting to the sponsor

SIMULATED · v1.0

Study / site	SART case ID / revision	Participant code or N/A	Sponsor safety case ID
GLENOS-01 / 014	SART-014-005 / r4	014-0005	GS-SAE-014-0003

D Reporting to the sponsor

Record the proof of reporting, not the content of the sponsor form. Every time limit comes from the CIP or the sponsor procedure. The PIR reads its safety reporting timeliness counts from this section.

Reporting step	Status	Record / reference
Reportable under the current CIP or sponsor procedure? (rule reference)	☑ Yes ☐ No ☐ Pending	CIP v4.2 section 9.4: SAEs and DDs that could have led to a SADE within 72 hours of awareness
Route used	☐ Sponsor portal ☐ eCRF ☑ SAE form ☑ DD form ☐ Emergency route ☐ Other	GLENOS EDC
Site awareness: date / time (from Section B)	28 Oct 2026, 08:05	
CIP reporting window, and due date / time	72 hours; due 31 Oct 2026, 08:05	
Initial notification sent: date / time, by whom, system reference	1 Nov 2026, 11:30; M. Silva (P-06); GS-SAE-014-0003 and GS-DD-014-0012	
Receipt confirmed / sponsor safety case ID	☑ Yes ☐ Pending	GS-SAE-014-0003
Initial notification on time?	☐ Yes ☑ No ☐ Pending ☐ N/A	Sent 27 hours late
First report complete?	☐ Complete ☑ Incomplete ☐ Pending ☐ N/A	Liner lot missing; sent 2 Nov after the explant label was read
If late or incomplete: where the delay arose	☐ Recognition ☑ Internal escalation ☐ PI assessment ☐ Access to documentation ☐ Transmission ☐ Other	Backup PI informed but not told the case was reportable; PI on leave
Correction made, and deviation record ID if the notification was late	Backup notification rule added to POT Section C; deviation PD-014-12; CAPA-014-2026-004	
Ethics committee or authority communication, only if the CIP assigns it to the site	☐ Sent ☑ Not required ☐ Pending	Not assigned to the site by the CIP

SART - Safety Assessment, Reporting and Tracking Template

Follow-up and new information

SIMULATED · v1.0

Study / site	SART case ID / revision	Participant code or N/A	Sheet number / total sheets
GLENOS-01 / 014	SART-014-005 / r4	014-0005	1 / 1

E Follow-up and new information

One line for each new piece of information or sponsor request. Add lines; never overwrite an earlier one.

No.	Date / time received	New clinical, technical or device information, and its source	Effect on the site assessment	Sponsor request or communication (reference)	Follow-up due	Sent / receipt reference	On time?	Owner (Role ID) and next step
1	29 Oct 2026, 09:10	Revision findings: liner separated, locking ring not engaged; liner kept (ICDL-014-005-R)	DD pathway opened; device relation confirmed	N/A	N/A	Added to the first report	☐ Yes ☐ No ☑ N/A	P-06: complete C2
2	3 Nov 2026, 14:00	Sponsor asks for the liner lot and the assembly steps (MM-Q-131)	None	MM-Q-131	6 Nov 2026	Answered 5 Nov; EDC receipt	☑ Yes ☐ No ☐ N/A	P-01: reply with ICDL and explant label
3	13 Nov 2026, 16:20	Sponsor determination: USADE; lot GS-L36-2607 quarantined at all sites	Anticipated status resolved	SAF-GS-0031	N/A	N/A	☐ Yes ☐ No ☑ N/A	P-01: quarantine stock; tell the team
4	19 Nov 2026, 10:00	Device analysis: locking ring out of tolerance (DA-0045)	Device cause confirmed; operator contribution reduced	DA-0045	N/A	N/A	☐ Yes ☐ No ☑ N/A	P-01: update the SCML context
5	9 Dec 2026, 11:00	6-week visit after revision: stable, no further dislocation (PFVR-014-005-R6)	Outcome: recovering	N/A	16 Dec 2026	Sent 10 Dec; EDC update	☑ Yes ☐ No ☐ N/A	P-06: 3-month visit, Jan 2027
6							☐ Yes ☐ No ☐ N/A	
7							☐ Yes ☐ No ☐ N/A	
8							☐ Yes ☐ No ☐ N/A	

If more lines are needed, reprint this sheet, number it and continue the numbering.

SART - Safety Assessment, Reporting and Tracking Template

Links and closure

SIMULATED · v1.0

Study / site	SART case ID / revision	Participant code or N/A	Local closure date
GLENOS-01 / 014	SART-014-005 / r4	014-0005	Open; next review 15 Jan 2027

F Records this case is sent to

Each line is a record that receives this case from the SART. Mark whether the link was made and write the receiving record's ID.

Record	When it applies	ID / date	Status
SCML	Every adverse event or device deficiency (DD) at the site · SCML event ID	SCML-014-004, 28 Oct 2026	☑ Linked ☐ Not required ☐ Pending
Deviation record	The initial notification was late; the sponsor's criteria decide its importance · Deviation ID	PD-014-12, 1 Nov 2026	☑ Linked ☐ Not required ☐ Pending
CSST, early review	Serious health threat, cluster, new category, or concentration in one component or surgeon · Review reference	Early CSST review, 4 Nov 2026 (serious event, new category)	☑ Linked ☐ Not required ☐ Pending
POT	The case, a delay or a pending decision needs the PI's oversight · POT review ID	POT-014-T-2026-11, triggered review	☑ Linked ☐ Not required ☐ Pending
SRR or CAPA	A recurring failure or a weak local control; not for a single case · Record ID	CAPA-014-2026-004: escalation while the PI is away	☑ Linked ☐ Not required ☐ Pending
PIR	Every reportable case enters the safety reporting timeliness count of its review period · PIR ID / period	PIR-014-2026-11, November 2026	☑ Linked ☐ Not required ☐ Pending
Sponsor device-deficiency route	The device deficiency needs return, analysis or action by the sponsor · Sponsor DD case ID	GS-DD-014-0012; RMA-0045	☑ Linked ☐ Not required ☐ Pending

G Closing the site's part of the case

Before closing, confirm that…	Result	Reference / limitation
Immediate protection and care documented (or N/A, with the reason)	☑ Yes ☐ No ☐ N/A	Emergency department and revision notes
Source references enough to reconstruct the case	☑ Yes ☐ No ☐ N/A	Section B source map
Current PI / site assessment documented	☑ Yes ☐ No ☐ N/A	C4, 3 Nov 2026
Sponsor submission and receipt documented	☑ Yes ☐ No ☐ N/A	GS-SAE-014-0003, GS-DD-014-0012
Timeliness and completeness known, or the limitation recorded	☑ Yes ☐ No ☐ N/A	Late and incomplete: D, PD-014-12
Follow-up completed, transferred or explicitly open	☑ Yes ☐ No ☐ N/A	Open: 3-month visit, Jan 2027
Sponsor / CEC determination recorded, or recorded as pending	☑ Yes ☐ No ☐ N/A	SAF-GS-0031, 13 Nov 2026
Links in Section F completed	☑ Yes ☐ No ☐ N/A	F, all seven linked

Local closure status

☐ Closed locally ☑ Open follow-up ☐ Transferred ☐ Not yet eligible

Current outcome and open items

Recovering after revision. Open: sponsor lot investigation and the 3-month visit.

Local closure does not close the sponsor, CEC, ethics committee or regulatory process.

Case reporter / initial site assessor

Name / role: Dr R. Teixeira / Sub-investigator, PI backup

Signature / date:

R. Teixeira 28 Oct 2026

PI / accountable investigator

Name / role: Dr Alex Morgan / PI

Signature / date:

A. Morgan 10 Dec 2026

Safety coordinator (if applicable)

Name / role: Maria Silva / Study Coordinator

Signature / date:

M. Silva 10 Dec 2026

Controlled location of the SART and linked system references: Investigator site file, tab 6 (Safety) / SART-014-005; GLENOS EDC GS-SAE-014-0003 and GS-DD-014-0012; SCML-014; RMA-0045