




cut along this line before use

Engineering Scout - calibration sheet (validation)

Same 20 mm reference square plus a printed test subject: a 40 x 8 mm bar.

Expected result: subject length 40.0 mm, width 8.0 mm (target |error| ≤ 0.5 mm at 640x480, 15-25 cm).

Use this page first. If the bar measures wrong, the problem is the rig (scale, focus, lighting), not the fastener.



= 100 mm on a ruler, or the print is scaled and the sheet is void



cut along this line before use

Engineering Scout - calibration sheet (working)

Reference: the solid square is 20 x 20 mm (engineering_scout_capture default --reference-mm 20).

Place ONE fastener on the white area, at least 20 mm from the square, not touching it. Nothing else in the camera's view.

Camera straight down, 15-25 cm above the sheet; light from the side so the fastener casts no hard shadow.

Matte paper. If the print looks grey or streaky, reprint - the square must be solid black.



= 100 mm on a ruler, or the print is scaled and the sheet is void