

MC QC FM01

Instrument No.	Id No./ Sr no	Description	Size	Qty.	Calibration	Acceptance Criteria	Max.Error in	Accepted	Calibration	Due Date	Location	Make	Certificate No	Calibration Agency	Remark
					Freq.		instrument	Yes/No	Date						
MIC/1	HG 263 / 485	2D DIGITAL HEIGHT GAUGE	0-625mm	1	ANNUALLY	± 0.05 mm	0.0036	Y	27-01-2026	26-01-2027	Q.C	MICROTECHNICA	IAP/HM/2026 /10401	MICROTECHNICA	
MIC/2	4-17120284	BEVEL PROTECTOR	0-360 °	1	ANNUALLY	± 10'	4.24 Min of Arc	Y	15-11-2025	14-11-2026	Q.C	ACCUPLUS	MH/CL/25/11/707/NABL	METHEAT SERVICES	Calibration frequency increased despite low usage.
MIC/3	8509	SURFACE PLATE	1600 X 1000 mm	1	ANNUALLY	± 0.05 mm	0.00750	Y	13-08-2025	12-08-2026	Q.C	MICRO FLAT	GMGS/N3994/2025	GM GAUGING SOLUTION	
MIC/4	756162	DEPTH MICROMETER	0-25 mm	1	ANNUALLY	± 0.025 mm	± 0.01	Y	05-10-2025	04-10-2026	Q.C	MITUTOYO	MH/CL/25/10/139/NABL	METHEAT SERVICES	
MIC/5	C-1147	EXTERNAL MICROMETER	25-50 mm	1	ANNUALLY	± 0.025 mm	±0.02	Y	27-09-2024	26-09-2025	Q.C	HIP	MH/CL/24/09/1181/NABL	METHEAT SERVICES	NOT IN USE
MIC/6	141198982	EXTERNAL DIGITAL MICROMETER	0-25mm	1	ANNUALLY	± 0.025 mm	±0.0041	Y	27-09-2025	26-09-2026	Q.C	BAKER	MH/CL/25/09/905/NABL	METHEAT SERVICES	
MIC/7	16225689	EXTERNAL MICROMETER	150-175 mm	1	ANNUALLY	± 0.025 mm	±0.02	Y	04-10-2025	03-10-2026	Q.C	MITUTOYO	MH/CL/25/10/187/NABL	METHEAT SERVICES	
MIC/8	8000947	EXTERNAL MICROMETER	150-300 mm	1	ANNUALLY	± 0.025 mm	±0.00828	Y	30-10-2025	29-10-2026	Q.C	MITUTOYO	GMGS/N5056/2025	GM GAUGING SOLUTION	
MIC/9	768731	BORE GAUGE	50-150 mm	1	ANNUALLY	± 0.025 mm	±0.00605	Y	30-10-2025	29-10-2026	Q.C	MITUTOYO	GMGS/N5059/2025	GM GAUGING SOLUTION	
MIC/10	1090966	BORE GAUGE	160-250 mm	1	ANNUALLY	± 0.025 mm	±0.0066	Y	29-10-2025	28-10-2026	Q.C	MITUTOYO	MH/CL/25/10/697/NABL	METHEAT SERVICES	
MIC/11	6403588	BORE GAUGE	18-35 mm	1	ANNUALLY	± 0.025 mm	±0.00605	Y	29-01-2026	28-01-2028	Q.C	MITUTOYO	MH/CL/26/01/1048/NABL	METHEAT SERVICES	
MIC/12	XQK321	PLUNGER TYPE DIAL INDICATOR	0-10 mm	1	ANNUALLY	± 0.025 mm	±0.0070	Y	27-09-2025	26-09-2026	Q.C	MITUTOYO	MH/CL/25/09/890/NABL	METHEAT SERVICES	
MIC/13	110443	SLIP GAUGE SET	1.0005-100mm	1 SET	ANNUALLY		Sheet Attached	Y	21-05-2025	21-05-2026	Q.C	M&W	CCS/25/S108/001	CREATIVE CALIBRATION SERVICES	
MIC/14	FF302523	DIGITAL VARNIER CALIPER	0-300mm	1	ANNUALLY	± 0.05 mm	±0.01873	Y	14-08-2025	13-08-2026	CNC	BAKER	GMGS/N4018/2025	GM GAUGING SOLUTION	
MIC/15	MG302428	DIGITAL VARNIER CALIPER	0-300mm	1	ANNUALLY	± 0.05 mm	±0.02	Y	31-03-2025	30-03-2026	CNC	BAKER	MH/CL/25/03/991/NABL	METHEAT SERVICES	Under Calibration / GM Gaueging
MIC/16	0122639	DIGITAL VARNIER CALIPER	0-300mm	1	ANNUALLY	± 0.05 mm	±0.02	Y	29-03-2024	28-03-2026	Q.C	MITUTOYO	MH/CL/24/03/716/NABL	MET-HEAT SERVICES	Under Calibration / GM Gaueging
MIC/17	BP-01	BEVEL PROTECTOR	0-90°	1	ANNUALLY	± 10'	6.7 Min of Arc	Y	30-10-2025	29-10-2026	Q.C	INOX	GMGS/N5057/2025	GM GAUGING SOLUTION	
MIC/18	HF302506	DIGITAL VARNIER CALIPER	0-300mm	1	ANNUALLY	± 0.05 mm	±0.02	Y	26-03-2026	25-03-2027	CNC	BAKER	GMGS/N1253/2026	GM GAUGING SOLUTION	
MIC/19	5113782	EXTERNAL MICROMETER	0-150 mm	1	ANNUALLY	± 0.025 mm	±0.0209	Y	02-07-2025	01-07-2026	QC	MITUTOYO	MH/CL/25/07/057/NABL	MET HEAT SERVICES	
MIC/20	APRG24	PLUNGER TYPE DIAL INDICATOR	0.001 - 1 mm	1	ANNUALLY		±0.0018	Y	02-07-2025	01-07-2026	Q.C	MITUTOYO	MH/CL/25/07/029/NABL	MET HEAT SERVICES	
MIC/21	975196	BORE GAUGE	10-18.5 mm	1	ANNUALLY	± 0.025 mm	±0.0066	Y	04-10-2025	03-10-2026	Q.C	MITUTOYO	MH/CL/25/10/108/NABL	MET HEAT SERVICES	
MIC/22	858368	BORE GAUGE	6-10 mm	1	ANNUALLY	± 0.025 mm	±0.0066	Y	27-09-2025	26-09-2026	QC	MITUTOYO	MH/CL/25/09/891/NABL	MET HEAT SERVICES	
MIC/23	9574640	BORE GAUGE	250-400 mm	1	ANNUALLY	± 0.025 mm	±0.0066	Y	29-10-2025	28-10-2026	QC	MITUTOYO	MH/CL/25/10/698/NABL	MET HEAT SERVICES	
MIC/24	180270421	EXTERNAL MICROMETER	300-400 mm	1	2 YRS	± 0.05 mm	±0.02	Y	16-04-2024	15-04-2026	QC	ACCUPLUS	MH/CL/24/04/345/NABL	METHEAT SERVICES	Calibration frequency increased despite low usage.

Microcraft Engineering Instruments List													MC QC FM01		
Instrument	Id No./ Sr no	Description	Size	Qty.	Calibration	Acceptance	Max.Error in	Accepted	Calibration	Due Date	Location	Make	Certificate No	Calibration Agency	Remark
No.					Freq.	Criteria	Instrument	Yes/No	Date						
MIC/24	180270421	EXTERNAL MICROMETER	300-400 mm	1	2 YRS	± 0.05 mm	±0.02	Y	16-04-2024	15-04-2026	QC	ACCUPLUS	MH/CL/24/04/345/NABL	METHEAT SERVICES	Calibration frequency increased despite low usage.
MIC/25	104-143 (61058509)	EXTERNAL MICROMETER	400-500 mm	1	2 YRS	± 0.05 mm		Y	NEW PURCHASE		QC	MITUTOYO			Calibration frequency increased despite low usage.
MIC/26	BBYF61	PLUNGER TYPE DIAL INDICATOR	0-10 mm	1	ANNUALLY	± 0.025 mm	±0.0073	Y	30-10-2025	29-10-2026	QC	MITUTOYO	GMGS/N5058/2025	GM GAUGING SOLUTION	
MIC/27	BCEL09	PLUNGER TYPE DIAL INDICATOR	0-10 mm	1	ANNUALLY	± 0.025 mm	±0.0073	Y	14-08-2025	13-08-2026	Q.C	MITUTOYO	MH/CL/24/08/021/NABL	MET HEAT SERVICES	
MIC/28	21793789	BORE GAUGE	50-150 mm	1	ANNUALLY	± 0.025 mm	±0.0066	Y	10-10-2025	09-10-2026	Q.C	MITUTOYO	MH/CL/25/10/297/NABL	MET HEAT SERVICES	
MIC/29	18575075	DIAL CALIPER	0-200 mm	1	ANNUALLY	± 0.05 mm	±0.026	Y	04-07-2025	03-07-2026	Q.C	MITUTOYO	MH/CL/25/07/115/NABL	MET HEAT SERVICES	
MIC/30	72291701	DIGITAL EXTERNAL MICROMETER	25-50 mm	1	ANNUALLY	± 0.025 mm	±0.00575	Y	02-07-2025	01-07-2026	Q.C	MITUTOYO	MH/CL/25/07/056/NABL	MET HEAT SERVICES	
MIC/31	KB302137	DIGITAL VARNIER CALIPER	0-300mm	1	ANNUALLY	± 0.05 mm	±0.02	Y	04-07-2025	03-07-2026	5 AXIS	BAKER	MH/CL/25/07/116/NABL	MET HEAT SERVICES	
MIC/32	72401284	EXTERNAL MICROMETER	0-150 MM	1	ANNUALLY	± 0.025 mm	±0.02		31-03-2025	30-03-2026	Q.C	MITUTOYO	MH/CL/25/03/1012/NABL	MET HEAT SERVICES	Under Calibration / MET HEAT
MIC/33	21752716	BORE GAUGE	35-60 mm	1	ANNUALLY	± 0.025 mm	±0.00605		14-08-2025	13-08-2026	Q.C	MITUTOYO	GMGS/N4014/2025	GM GAUGING	
MIC/34	FL05128	DIGITAL VARNIER CALIPER	0-300 mm	1	ANNUALLY	± 0.05 mm	± 0.02	Y	28-06-2024	27-06-2025		BAKER	MH/CL/24/06/1493/NABL	MET HEAT	MISPLACED
MIC/35	1512A	RADIUS GAUGE	1-7 MM	1	ANNUALLY		± 0.02047	Y	02-07-2025	01-07-2026	Q.C	KRISTEL	MH/CL/25/07/058/NABL	MET HEAT SERVICES	
MIC/36	QER886	PLUNGER TYPE DIAL GAUGE	0-10 MM	1	ANNUALLY	± 0.025 mm	0.0007	Y	23-10-2024	22-10-2025	Q.C	MITUTOYO	MH/CL/24/10/653/NABL	MET HEAT	NOT IN USED
MIC/37	LF306574	DIGITAL VARNIER CALIPER	0-300 mm	1	ANNUALLY	± 0.05 mm	± 0.02	Y	29-03-2025	28-03-2026	QC	BAKER	MH/CL/25/03/859/NABL	MET HEAT SERVICES	Under Calibration / MET HEAT
MIC/38	MG302547	DIGITAL VARNIER CALIPER	0-300 mm	1	ANNUALLY	± 0.05 mm	± 0.02	Y	23-03-2026	22-03-2027	VMC	BAKER	MH/CL/26/03/717/NABL	MET HEAT SERVICES	
MIC/39	231095120	EXTERNAL DIGITAL MICROMETER	0-25mm	1	ANNUALLY	± 0.02 mm	0.0057	Y	13-11-2025	12-11-2026	QC	BAKER	MH/CL/25/11/656/NABL	MET HEAT SERVICES	
MIC/40	NJ307189	DIGITAL VARNIER CALIPER	0-300 mm	1	ANNUALLY	± 0.05 mm	± 0.02	Y	02-01-2026	01-01-2027	QC	BAKER	MH/CL/26/01/021/NABL	MET HEAT SERVICES	
MIC/41	NB307550	DIGITAL VARNIER CALIPER	0-300 mm	1	ANNUALLY	± 0.05 mm	± 0.02 mm	Y	02-01-2026	01-01-2027	VMC	BAKER	MH/CL/26/01/020/NABL	MET HEAT SERVICES	
MIC/42	459103	MICRO BORE GAUGE	7-10 mm	1	ANNUALLY				NEW PURCHASE		QC	MITUTOYO			
MIC/43	AF-1557	EXTERNAL BLADE MICROMETER	0-25mm	1	ANNUALLY				NEW PURCHASE		QC	BAKER			

